

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS051921\
 Data File : PS015038.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2021 18:58
 Operator : DD\AJ
 Sample : M2409-01MS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 BACKFILL-ASHCROFT-REAR-06MS

Manual Integrations
 APPROVED

Ankita
 5/20/2021 1:44:38 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 04:17:37 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS051021.M
 Quant Title : 8080.M
 QLast Update : Tue May 11 00:04:04 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.155	6.884	587.5E6	183.6E6	351.116	379.313m
Target Compounds							
1) T	Dalapon	3.003	2.602	615.0E6	269.8E6	300.661	399.184m#
2) T	3,5-DICHL...	6.382	6.049	700.6E6	261.6E6	298.897	338.641m
3) T	4-Nitroph...	6.993	6.507	260.8E6	79381198	313.702	267.284
5) T	DICAMBA	7.342	7.048	1969.6E6	689.4E6	300.590	321.903
6) T	MCPD	7.504	7.139	149.8E6	54788972	30.639	30.999
7) T	MCPA	7.652	7.352	179.1E6	103.7E6	25.777	41.300 #
8) T	DICHLORPROP	8.029	7.675	469.0E6	172.5E6	294.071	318.929
9) T	2,4-D	8.262	7.974	643.8E6	229.6E6	361.798	377.085
10) T	Pentachlo...	8.562	8.393	6777.9E6	2243.8E6	288.288	311.197
11) T	2,4,5-TP ...	9.106	8.754	2899.8E6	992.2E6	317.299	344.280
12) T	2,4,5-T	9.394	9.138	2754.9E6	861.4E6	301.237	322.773
13) T	2,4-DB	9.943	9.643	342.2E6	100.1E6	197.226	283.691 #
14) T	DINOSEB	11.087	9.970	762.3E6	222.2E6	106.299	108.498
15) T	Picloram	10.921	10.958	3534.0E6	1278.2E6	290.929	320.777
16) T	DCPA	11.354	10.893	3548.4E6	1359.5E6	259.777	318.952

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS051921\
Data File : PS015038.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2021 18:58
Operator : DD\AJ
Sample : M2409-01MS
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
BACKFILL-ASHCROFT-REAR-06MS

Manual Integrations
APPROVED

Ankita
5/20/2021 1:44:38 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 04:17:37 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS051021.M
Quant Title : 8080.M
QLast Update : Tue May 11 00:04:04 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

