

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS052021\
 Data File : PS015051.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 May 2021 14:02
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 15:00:23 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS051021.M
 Quant Title : 8080.M
 QLast Update : Tue May 11 00:04:04 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.153	6.881	1207.4E6	366.2E6	721.594	756.789
Target Compounds							
1) T	Dalapon	3.004	2.603	1262.2E6	435.1E6	617.094	643.676
2) T	3,5-DICHL...	6.380	6.045	1565.8E6	545.6E6	668.071	706.288
3) T	4-Nitroph...	6.988	6.499	546.1E6	210.5E6	656.777	708.820
5) T	DICAMBA	7.340	7.047	4539.4E6	1552.8E6	692.767	725.024
6) T	MCPD	7.507	7.142	344.9E6	127.8E6	70.538	72.297
7) T	MCPA	7.655	7.357	472.4E6	176.2E6	67.996	70.134
8) T	DICHLORPROP	8.026	7.671	1129.1E6	396.3E6	707.936	732.824
9) T	2,4-D	8.256	7.968	1301.3E6	476.2E6	731.257	781.953
10) T	Pentachlo...	8.560	8.390	16123.1E6	5153.1E6	685.775	714.686
11) T	2,4,5-TP ...	9.104	8.752	6488.3E6	2151.7E6	709.948	746.576
12) T	2,4,5-T	9.389	9.131	6351.1E6	2020.1E6	694.463	756.900
13) T	2,4-DB	9.936	9.637	1149.4E6	266.8E6	662.422	756.350
14) T	DINOSEB	11.085	9.965	4908.0E6	1517.1E6	684.425	740.677
15) T	Picloram	10.913	10.952	8153.3E6	2773.3E6	671.204	695.988
16) T	DCPA	11.351	10.891	8344.3E6	2787.5E6	610.881	653.994

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS052021\
Data File : PS015051.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 May 2021 14:02
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 15:00:23 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS051021.M
Quant Title : 8080.M
QLast Update : Tue May 11 00:04:04 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

