

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS052022\
 Data File : PS019494.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 May 2022 13:37
 Operator : AJ\MA
 Sample : PB144929BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB144929BS

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 05/21/2022
 Supervised By : Ankita Jodhani 05/23/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 22:39:04 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS050422.M
 Quant Title : 8080.M
 QLast Update : Wed May 04 16:27:11 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.220	6.798	1454.7E6	380.9E6	532.696	491.707
Target Compounds							
1) T	Dalapon	3.027	2.535	2225.7E6	437.6E6	461.037	492.325
2) T	3,5-DICHL...	6.430	5.975	1639.4E6	568.9E6	463.817	487.838
3) T	4-Nitroph...	7.062	6.420	486.0E6	172.6E6	501.525	489.217
5) T	DICAMBA	7.408	6.962	5112.8E6	1683.5E6	462.937	502.402
6) T	MCPD	7.576	7.057	381.6E6	128.8E6	42.686	48.429
7) T	MCPA	7.725	7.269	520.9E6	178.2E6	41.795	47.851
8) T	DICHLORPROP	8.103	7.585	1276.9E6	417.3E6	456.435	486.888
9) T	2,4-D	8.338	7.881	1402.7E6	487.3E6	444.639	487.942
10) T	Pentachlo...	8.640	8.304	19194.4E6	5963.6E6	488.575	530.503
11) T	2,4,5-TP ...	9.194	8.667	8138.4E6	2461.7E6	450.256	511.509
12) T	2,4,5-T	9.482	9.047	9736.8E6	2148.3E6	460.344	496.144
13) T	2,4-DB	10.039	9.554	1657.1E6	256.2E6	386.432	450.577
14) T	DINOSEB	11.186	9.875	8898.6E6	1622.2E6	538.937	559.798
15) T	Picloram	11.044	10.881	10304.1E6	2774.9E6	352.517m	463.974 #
16) T	DCPA	11.454	10.805	12387.8E6	3048.5E6	473.340	519.296

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS052022\
Data File : PS019494.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 May 2022 13:37
Operator : AJ\MA
Sample : PB144929BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB144929BS

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 05/21/2022
Supervised By : Ankita Jodhani 05/23/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 22:39:04 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS050422.M
Quant Title : 8080.M
QLast Update : Wed May 04 16:27:11 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

