

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS052024\
 Data File : PS026841.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 May 2024 03:08
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 21 04:50:30 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS051424.M
 Quant Title : 8080.M
 QLast Update : Wed May 15 02:07:16 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.254	7.773	1812.7E6	1201.9E6	671.772	658.586
Target Compounds							
1) T	Dalapon	2.648	2.712	2386.4E6	2053.3E6	609.307	607.905
2) T	3,5-DICHL...	6.422	6.722	2394.9E6	1674.3E6	642.931	617.837
3) T	4-Nitroph...	7.054	7.300	1109.9E6	772.1E6	616.838	576.650
5) T	DICAMBA	7.440	7.974	7093.8E6	5113.0E6	650.135	631.231
6) T	MCPD	7.621	8.076	524.6E6	339.6E6	65.421	62.686
7) T	MCPA	7.769	8.320	737.8E6	503.9E6	64.335	64.819
8) T	DICHLORPROP	8.143	8.690	2016.0E6	1362.7E6	632.823	652.143
9) T	2,4-D	8.373	9.020	2277.4E6	1484.8E6	639.166	653.546
10) T	Pentachlo...	8.670	9.551	29445.9E6	20252.5E6	679.295	672.513
11) T	2,4,5-TP ...	9.247	9.925	11367.6E6	8009.2E6	650.703	672.101
12) T	2,4,5-T	9.538	10.346	11926.9E6	8033.4E6	651.851	668.936
13) T	2,4-DB	10.111	10.911	1611.1E6	1013.5E6	628.374	658.017
14) T	DINOSEB	11.319	11.291	7688.9E6	5381.7E6	645.830m	654.055
15) T	Picloram	11.128	12.388	15547.6E6	11844.7E6	658.297	664.980
16) T	DCPA	11.611	12.334	13567.2E6	9748.9E6	666.874	674.792

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS052024\
Data File : PS026841.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 May 2024 03:08
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 21 04:50:30 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS051424.M
Quant Title : 8080.M
QLast Update : Wed May 15 02:07:16 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

