

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS052025\
 Data File : PS030306.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 May 2025 16:24
 Operator : AR\AJ
 Sample : PB168079BSD
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB168079BSD

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 05/21/2025
 Supervised By : mohammad ahmed 05/22/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 21 02:16:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS051225.M
 Quant Title : 8080.M
 QLast Update : Mon May 12 14:29:24 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.930	7.449	1566.9E6	423.4E6	550.211	529.097
Target Compounds							
1) T	Dalapon	2.445	2.519	2430.6E6	965.7E6	494.484	475.488
2) T	3,5-DICHL...	6.136	6.449	2130.9E6	548.3E6	510.942	470.268
3) T	4-Nitroph...	6.721	6.984	1033.0E6	456.9E6	500.072	437.643
5) T	DICAMBA	7.106	7.634	6019.8E6	2357.3E6	520.859	498.593
6) T	MCPD	7.284	7.741	382.5E6	90172413	52.518	49.013
7) T	MCPA	7.426	7.970	511.7E6	129.2E6	49.291	48.966
8) T	DICHLORPROP	7.786	8.326	1515.9E6	588.6E6	518.922	498.195
9) T	2,4-D	8.006	8.638	1704.9E6	658.2E6	519.949	510.009
10) T	Pentachlo...	8.283	9.137	21950.7E6	12973.2E6	541.982	525.550
11) T	2,4,5-TP ...	8.846	9.515	8602.3E6	5093.8E6	531.140	517.114
12) T	2,4,5-T	9.128	9.919	8792.1E6	4703.4E6	532.365	511.043
13) T	2,4-DB	9.685	10.477	1444.0E6	444.3E6	553.314	447.754
14) T	DINOSEB	10.845	10.850	5959.5E6	3447.6E6	521.097	504.620
15) T	Picloram	10.666	11.888	11231.8E6	6167.4E6	525.171	438.240m
16) T	DCPA	11.147	11.885	10736.7E6	7638.4E6	539.220	565.364m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS052025\
Data File : PS030306.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 May 2025 16:24
Operator : AR\AJ
Sample : PB168079BSD
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB168079BSD

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 05/21/2025
Supervised By : mohammad ahmed 05/22/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 21 02:16:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS051225.M
Quant Title : 8080.M
QLast Update : Mon May 12 14:29:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

