

Data Path : P:\HPCHEM1\ECD_S\Data\PS052118\
 Data File : PS000128.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 May 2018 17:17
 Operator : UA
 Sample : HSTDICC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC500

Manual Integrations
 APPROVED

UGO
 5/22/2018 4:38:44 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 22 01:50:44 2018
 Quant Method : P:\HPCHEM1\ECD_S\Method\PS052118.M
 Quant Title : 8080.M
 QLast Update : Tue May 22 01:38:24 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.825	11.195	177.8E6	158.0E6	512.557	511.593
Target Compounds							
1) T	Dalapon	3.169	2.549	217.0E6	180.0E6	458.696	458.762
2) T	3,5-DICHL...	10.237	9.338	253.1E6	207.3E6	482.450	476.404
3) T	4-Nitroph...	11.422	10.342	107.7E6	85700788	475.508	473.091
5) T	DICAMBA	12.145	11.485	627.8E6	523.8E6	486.589	474.929
6) T	MCPD	12.482	11.707	52346668	43407227	46.077m	46.457m
7) T	MCPA	12.734	12.084	88488410	69008386	48.233	47.168m
8) T	DICHLORPROP	13.338	12.608	164.8E6	136.6E6	486.679	479.507
9) T	2,4-D	13.696	13.080	214.9E6	181.5E6	483.318	476.536
10) T	Pentachlo...	14.096	13.644	2496.0E6	2173.2E6	513.876	501.082
11) T	2,4,5-TP ...	14.963	14.252	972.8E6	834.0E6	499.272	487.464
12) T	2,4,5-T	15.360	14.785	961.3E6	799.3E6	496.099	485.277
13) T	2,4-DB	16.116	15.472	129.5E6	99842272	479.101	471.033
14) T	DINOSEB	17.629	15.905	703.7E6	588.1E6	487.096	477.225
15) T	Picloram	17.385	17.151	1272.8E6	1026.2E6	490.245	476.426
16) T	DCPA	17.969	17.092	1215.5E6	1022.3E6	512.979	499.212

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : P:\HPCHEM1\ECD_S\Data\PS052118\
Data File : PS000128.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 May 2018 17:17
Operator : UA
Sample : HSTDICC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDICC500

Manual Integrations
APPROVED

UGO
5/22/2018 4:38:44 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 22 01:50:44 2018
Quant Method : P:\HPCHEM1\ECD_S\Method\PS052118.M
Quant Title : 8080.M
QLast Update : Tue May 22 01:38:24 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

