

Method Path : P:\HPCHEM1\ECD_S\Method\
 Method File : PS052118.M
 Title : 8080.M
 Last Update : Tue May 22 01:55:55 2018
 Response Via : Initial Calibration

Calibration Files

750 =PS000129.D 1000 =PS000130.D 1500 =PS000131.D
 500 =PS000128.D 200 =PS000127.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	4.722	4.713	4.718	4.769	4.978	4.780	E5	2.36
2) T	3,5-DICHLOROBENZOI...	5.317	5.195	5.031	5.444	5.757	5.349	E5	5.13
3) T	4-Nitrophenol	2.297	2.226	2.168	2.366	2.590	2.329	E5	7.03
4) S	2,4-DCAA	3.529	3.474	3.316	3.556	3.779	3.531	E5	4.73
5) T	DICAMBA	1.309	1.282	1.234	1.336	1.379	1.308	E6	4.17
6) T	MCPP	1.132	1.141	1.158	1.114	1.036	1.116	E6	4.26
7) T	MCPA	1.896	1.753	1.787	1.903	2.031	1.874	E6	5.86
8) T	DICHLORPROP	3.426	3.322	3.289	3.506	3.710	3.450	E5	4.89
9) T	2,4-D	4.490	4.413	4.310	4.572	4.764	4.510	E5	3.82
10) T	Pentachlorophenol	4.996	4.773	4.405	5.255	5.562	4.998	E6	8.87
11) T	2,4,5-TP (SILVEX)	1.986	1.926	1.833	2.048	2.117	1.982	E6	5.52
12) T	2,4,5-T	1.974	1.922	1.831	2.024	2.053	1.961	E6	4.49
13) T	2,4-DB	2.719	2.676	2.690	2.726	2.783	2.719	E5	1.51
14) T	DINOSEB	1.468	1.434	1.380	1.497	1.538	1.463	E6	4.13
15) T	Picloram	2.643	2.592	2.470	2.680	2.694	2.616	E6	3.46
16) T	DCPA	2.426	2.336	2.183	2.532	2.628	2.421	E6	7.12

Signal #2 Calibration Files

750 =PS000129.D 1000 =PS000130.D 1500 =PS000131.D
 500 =PS000128.D 200 =PS000127.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	3.923	3.911	3.907	3.957	4.052	3.950	E5	1.52
2) T	3,5-DICHLOROBENZOI...	4.386	4.329	4.230	4.457	4.674	4.415	E5	3.78
3) T	4-Nitrophenol	1.825	1.791	1.746	1.884	2.066	1.862	E5	6.67
4) S	2,4-DCAA	3.116	3.028	3.048	3.159	3.354	3.141	E5	4.14
5) T	DICAMBA	1.113	1.105	1.080	1.115	1.120	1.106	E6	1.43
6) T	MCPP	9.329	9.315	9.483	9.236	8.913	9.255	E5	2.28
7) T	MCPA	1.461	1.452	1.449	1.484	1.578	1.485	E6	3.65
8) T	DICHLORPROP	2.840	2.845	2.804	2.907	3.049	2.889	E5	3.35
9) T	2,4-D	3.832	3.785	3.754	3.861	3.990	3.844	E5	2.38
10) T	Pentachlorophenol	4.441	4.298	4.034	4.575	4.678	4.405	E6	5.72
11) T	2,4,5-TP (SILVEX)	1.730	1.708	1.650	1.756	1.757	1.720	E6	2.57
12) T	2,4,5-T	1.663	1.648	1.596	1.683	1.652	1.648	E6	1.97
13) T	2,4-DB	2.117	2.122	2.137	2.102	2.128	2.121	E5	0.61
14) T	DINOSEB	1.247	1.223	1.209	1.251	1.255	1.237	E6	1.63
15) T	Picloram	2.193	2.164	2.098	2.160	2.037	2.131	E6	2.94
16) T	DCPA	2.085	2.037	1.940	2.130	2.142	2.067	E6	3.97

(#) = Out of Range