

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS052119\
 Data File : PS004662.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 May 2019 09:01
 Operator : EI\SJ
 Sample : HSTDICV500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICV500

Manual Integrations
 APPROVED

Ankita
 5/22/2019 9:40:02 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 21 11:48:23 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052019.M
 Quant Title : 8080.M
 QLast Update : Tue May 21 04:53:35 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.595	5.301	1022.6E6	720.1E6	729.549	746.333
Target Compounds							
1) T	Dalapon	2.116	1.811	1165.3E6	894.2E6	702.982m	692.921
2) T	3,5-DICHL...	4.941	4.593	1205.9E6	853.6E6	699.236	692.795
3) T	4-Nitroph...	5.441	4.976	425.0E6	322.7E6	681.600	680.893
5) T	DICAMBA	5.756	5.436	3667.7E6	2591.8E6	690.756	701.591
6) T	MCPD	5.916	5.525	291.2E6	195.4E6	69.948	71.307
7) T	MCPA	6.053	5.711	402.3E6	281.2E6	69.450	68.721
8) T	DICHLORPROP	6.396	5.992	900.3E6	656.5E6	687.958	696.347
9) T	2,4-D	6.611	6.264	964.6E6	754.9E6	705.198	698.398
10) T	Pentachlo...	6.883	6.642	12264.3E6	8661.1E6	698.394	705.017
11) T	2,4,5-TP ...	7.420	7.005	5078.9E6	3646.4E6	693.316	708.676
12) T	2,4,5-T	7.694	7.403	5178.3E6	3198.0E6	695.310	717.570
13) T	2,4-DB	8.226	7.855	679.3E6	465.1E6	703.257	711.606
14) T	DINOSEB	9.329	8.171	4214.0E6	2701.4E6	712.423	702.731
15) T	Picloram	9.155	9.111	7527.4E6	5332.8E6	722.398	714.800
16) T	DCPA	9.600	9.071	6484.0E6	4620.2E6	728.041	713.211

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS052119\
Data File : PS004662.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 May 2019 09:01
Operator : EI\SJ
Sample : HSTDICV500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICV500

Manual Integrations
APPROVED

Ankita
5/22/2019 9:40:02 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 21 11:48:23 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052019.M
Quant Title : 8080.M
QLast Update : Tue May 21 04:53:35 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

