

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS052119\
 Data File : PS004671.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 May 2019 12:32
 Operator : EI\SJ
 Sample : K2905-01MS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 T-3MS

Manual Integrations
 APPROVED

Ankita
 5/22/2019 9:40:14 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 22 01:55:11 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052019.M
 Quant Title : 8080.M
 QLast Update : Tue May 21 04:53:35 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.576	5.285	434.1E6	285.5E6	309.690	295.953
Target Compounds							
1) T	Dalapon	2.098	1.795	254.2E6	113.0E6	153.323m	87.567m#
2) T	3,5-DICHL...	4.924	4.578	499.5E6	355.1E6	289.640	288.202m
3) T	4-Nitroph...	5.423	4.961	140.6E6	129.6E6	225.527m	273.557
5) T	DICAMBA	5.736	5.419	1204.0E6	852.9E6	226.753	230.880m
6) T	MCPD	5.892	5.505	125.7E6	76849570	30.187m	28.047
7) T	MCPA	6.029	5.688	136.2E6	131.4E6	23.511	32.115 #
8) T	DICHLORPROP	6.377	5.976	375.0E6	275.7E6	286.523	292.397
9) T	2,4-D	6.593	6.248	436.5E6	367.1E6	319.125	339.674
10) T	Pentachlo...	6.864	6.626	5466.0E6	3911.2E6	311.262	318.373
11) T	2,4,5-TP ...	7.401	6.988	2115.5E6	1626.5E6	288.785	316.118
12) T	2,4,5-T	7.675	7.387	2109.8E6	1030.2E6	283.294	231.152
13) T	2,4-DB	8.208	7.839	313.6E6	188.1E6	324.665	287.794
14) T	DINOSEB	9.311	8.153	397.8E6	337.0E6	67.258m	87.656 #
15) T	Picloram	9.141	9.097	1518.2E6	957.4E6	145.696	128.326
16) T	DCPA	9.581	9.054	449.7E6	190.1E6	50.493m	29.351m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS052119\
Data File : PS004671.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 May 2019 12:32
Operator : EI\SJ
Sample : K2905-01MS
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
T-3MS

Manual Integrations
APPROVED

Ankita
5/22/2019 9:40:14 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 22 01:55:11 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052019.M
Quant Title : 8080.M
QLast Update : Tue May 21 04:53:35 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

