

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS052119\
 Data File : PS004677.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 May 2019 14:44
 Operator : EI\SJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 22 01:57:38 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052019.M
 Quant Title : 8080.M
 QLast Update : Tue May 21 04:53:35 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.575	5.284	1029.7E6	727.8E6	734.550	754.336
Target Compounds							
1) T	Dalapon	2.099	1.795	1137.6E6	804.6E6	686.270	623.518
2) T	3,5-DICHL...	4.923	4.577	1225.6E6	866.0E6	710.674	702.817
3) T	4-Nitroph...	5.421	4.959	436.8E6	327.1E6	700.563	690.151
5) T	DICAMBA	5.736	5.418	3741.8E6	2620.0E6	704.714	709.216
6) T	MCPD	5.896	5.508	298.6E6	197.9E6	71.735	72.240
7) T	MCPA	6.033	5.693	414.4E6	285.9E6	71.543	69.863
8) T	DICHLORPROP	6.376	5.975	916.9E6	663.3E6	700.597	703.572
9) T	2,4-D	6.590	6.247	977.5E6	763.7E6	714.583	706.535
10) T	Pentachlo...	6.863	6.625	12574.1E6	8727.9E6	716.032	710.462
11) T	2,4,5-TP ...	7.400	6.987	5354.4E6	3682.3E6	730.922	715.662
12) T	2,4,5-T	7.674	7.385	5381.7E6	3161.4E6	722.618	709.346
13) T	2,4-DB	8.206	7.837	746.9E6	472.2E6	773.265	722.507
14) T	DINOSEB	9.310	8.153	4239.7E6	2733.4E6	716.773	711.045
15) T	Picloram	9.135	9.093	7382.1E6	5156.6E6	708.454	691.192
16) T	DCPA	9.581	9.053	6580.7E6	4668.3E6	738.892	720.632

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS052119\
Data File : PS004677.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 May 2019 14:44
Operator : EI\SJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 22 01:57:38 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052019.M
Quant Title : 8080.M
QLast Update : Tue May 21 04:53:35 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

