

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS052119\
 Data File : PS004701.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 May 2019 00:31
 Operator : EI\SJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 22 02:08:31 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052019.M
 Quant Title : 8080.M
 QLast Update : Tue May 21 04:53:35 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.583	5.292	1051.0E6	720.4E6	749.784	746.685
Target Compounds							
1) T	Dalapon	2.109	1.804	1086.8E6	783.4E6	655.643	607.042
2) T	3,5-DICHL...	4.931	4.585	1245.3E6	859.1E6	722.120	697.269
3) T	4-Nitroph...	5.429	4.967	437.6E6	322.6E6	701.901	680.664
5) T	DICAMBA	5.744	5.426	3801.7E6	2598.1E6	715.995	703.289
6) T	MCPD	5.903	5.516	306.2E6	196.3E6	73.559	71.645
7) T	MCPA	6.041	5.701	421.2E6	283.3E6	72.715	69.236
8) T	DICHLORPROP	6.383	5.982	950.9E6	659.1E6	726.625	699.080
9) T	2,4-D	6.599	6.254	990.8E6	749.2E6	724.313	693.117
10) T	Pentachlo...	6.871	6.632	12713.4E6	8694.0E6	723.968	707.696
11) T	2,4,5-TP ...	7.408	6.994	5334.8E6	3641.9E6	728.243	707.805
12) T	2,4,5-T	7.681	7.393	5294.3E6	3108.1E6	710.879	697.384
13) T	2,4-DB	8.214	7.845	735.9E6	464.7E6	761.833	710.891
14) T	DINOSEB	9.316	8.161	4168.8E6	2664.0E6	704.786	693.015
15) T	Picloram	9.143	9.101	7306.4E6	5053.3E6	701.188	677.340
16) T	DCPA	9.587	9.061	6540.4E6	4632.0E6	734.370	715.035

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS052119\
Data File : PS004701.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 May 2019 00:31
Operator : EI\SJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 22 02:08:31 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052019.M
Quant Title : 8080.M
QLast Update : Tue May 21 04:53:35 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

