

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS052422\
 Data File : PS019510.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 May 2022 21:57
 Operator : AJ\MA
 Sample : HSTDICC1500
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC1500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 25 01:58:07 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052422.M
 Quant Title : 8080.M
 QLast Update : Wed May 25 01:57:58 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.198	6.782	3524.8E6	1172.6E6	1489.276	1518.065
Target Compounds							
1) T	Dalapon	3.016	2.529	6791.4E6	1369.4E6	1363.765	1376.777
2) T	3,5-DICHL...	6.412	5.962	4194.1E6	1638.7E6	1366.280	1385.837
3) T	4-Nitroph...	7.027	6.398	1531.9E6	588.8E6	1429.898	1375.924
5) T	DICAMBA	7.388	6.947	12560.4E6	4638.0E6	1365.891	1383.334
6) T	MCP	7.564	7.052	1124.8E6	397.1E6	144.427	143.989
7) T	MCPA	7.714	7.266	1548.2E6	547.1E6	141.055	141.449
8) T	DICHLORPROP	8.079	7.566	3451.5E6	1240.7E6	1388.440	1416.788
9) T	2,4-D	8.307	7.857	4074.6E6	1501.7E6	1409.889	1413.838
10) T	Pentachlo...	8.615	8.284	40143.6E6	14093.8E6	1285.987	1317.353
11) T	2,4,5-TP ...	9.164	8.644	20314.5E6	6710.6E6	1371.749	1381.052
12) T	2,4,5-T	9.449	9.021	23924.6E6	6409.2E6	1358.951	1390.503
13) T	2,4-DB	10.000	9.525	5912.2E6	913.3E6	1442.948	1454.815
14) T	DINOSEB	11.156	9.852	19849.8E6	4863.6E6	1360.253	1390.062
15) T	Picloram	10.977	10.830	39630.3E6	8862.6E6	1381.419	1398.609
16) T	DCPA	11.423	10.781	30271.8E6	8106.7E6	1362.404	1382.377

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS052422\
Data File : PS019510.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 May 2022 21:57
Operator : AJ\MA
Sample : HSTDICC1500
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC1500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 25 01:58:07 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052422.M
Quant Title : 8080.M
QLast Update : Wed May 25 01:57:58 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

