

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS052519\
 Data File : PS004778.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 May 2019 16:17
 Operator : SM\AJ
 Sample : PB119997BS
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB119997BS

Manual Integrations
 APPROVED

Ankita
 5/28/2019 11:02:25 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 27 00:55:01 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052019.M
 Quant Title : 8080.M
 QLast Update : Tue May 21 04:53:35 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.593	5.313	841.4E6	603.2E6	600.267	625.199m
Target Compounds							
1) T	Dalapon	2.119	1.828	805.4E6	488.8E6	485.836	378.769
2) T	3,5-DICHL...	4.941	4.606	952.3E6	645.9E6	552.197	524.189
3) T	4-Nitroph...	5.439	4.988	311.4E6	241.4E6	499.363	509.402
5) T	DICAMBA	5.754	5.447	2617.3E6	1832.1E6	492.941	495.943
6) T	MCPD	5.912	5.535	230.2E6	143.9E6	55.289	52.506
7) T	MCPA	6.049	5.719	286.0E6	211.1E6	49.372m	51.600
8) T	DICHLORPROP	6.393	6.003	675.3E6	491.2E6	515.968	521.018
9) T	2,4-D	6.608	6.275	815.6E6	648.0E6	596.274	599.540
10) T	Pentachlo...	6.880	6.652	10091.4E6	6949.9E6	574.658	565.727
11) T	2,4,5-TP ...	7.417	7.015	4021.6E6	2918.8E6	548.979	567.275
12) T	2,4,5-T	7.690	7.413	3818.1E6	2364.1E6	512.661	530.458
13) T	2,4-DB	8.224	7.866	572.8E6	450.9E6	593.000	689.778
14) T	DINOSB	9.325	8.181	2948.8E6	1982.3E6	498.532	515.676
15) T	Picloram	9.152	9.123	4599.9E6	3204.8E6	441.452	429.570
16) T	DCPA	9.597	9.082	3299.5E6	2072.9E6	370.477	319.992

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS052519\
Data File : PS004778.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 May 2019 16:17
Operator : SM\AJ
Sample : PB119997BS
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
PB119997BS

Manual Integrations
APPROVED

Ankita
5/28/2019 11:02:25 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 27 00:55:01 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052019.M
Quant Title : 8080.M
QLast Update : Tue May 21 04:53:35 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

