

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS052720\
 Data File : PS010417.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 May 2020 09:42
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 5/28/2020 9:58:18 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 28 15:58:40 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS051120.M
 Quant Title : 8080.M
 QLast Update : Mon May 11 17:06:53 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.234	6.919	579.2E6	415.5E6	823.837	774.792
Target Compounds							
1) T	Dalapon	3.044	2.624	641.2E6	531.3E6	719.329	655.769
2) T	3,5-DICHL...	6.446	6.068	745.2E6	607.1E6	772.443	745.434
3) T	4-Nitroph...	7.064	6.523	241.3E6	237.1E6	729.552	713.535
5) T	DICAMBA	7.427	7.089	2176.4E6	1645.9E6	797.855	751.707
6) T	MCPD	7.597	7.187	170.3E6	140.1E6	77.784	76.332
7) T	MCPA	7.747	7.403	234.0E6	194.7E6	75.108	71.585
8) T	DICHLORPROP	8.120	7.715	515.9E6	414.1E6	772.802	733.540
9) T	2,4-D	8.349	8.009	531.1E6	492.5E6	751.768m	710.233
10) T	Pentachlo...	8.661	8.440	7398.4E6	5261.7E6	813.645	776.587
11) T	2,4,5-TP ...	9.211	8.801	2953.0E6	2216.6E6	801.251	765.186
12) T	2,4,5-T	9.496	9.180	2770.8E6	2211.2E6	787.428	760.520
13) T	2,4-DB	10.049	9.686	365.0E6	294.6E6	725.598	739.149
14) T	DINOSEB	11.212	10.020	2205.8E6	1657.4E6	805.033	777.788
15) T	Picloram	11.028	11.001	3279.0E6	2959.0E6	715.187	700.863
16) T	DCPA	11.477	10.949	3623.0E6	2779.0E6	800.922	777.393

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS052720\
Data File : PS010417.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 May 2020 09:42
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
5/28/2020 9:58:18 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 28 15:58:40 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS051120.M
Quant Title : 8080.M
QLast Update : Mon May 11 17:06:53 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

