

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS052721\
 Data File : PS015092.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 May 2021 21:44
 Operator : DD\AJ
 Sample : HSTDICV750
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 28 03:45:00 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052721.M
 Quant Title : 8080.M
 QLast Update : Fri May 28 03:42:14 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	Dalapon	2.288	2.333 E6	-2.0	98	0.00
2 T	3,5-DICHLORO BENZOIC ACID	2.800	2.822 E6	-0.8	99	0.00
3 T	4-Nitrophenol	1.196	1.209 E6	-1.1	99	0.00
4 S	2,4-DCAA	2.113	2.108 E6	0.2	98	0.00
5 T	DICAMBA	7.927	8.113 E6	-2.3	99	0.00
6 T	MCP P	6.114	6.273 E6	-2.6	97	0.00
7 T	MCPA	8.765	8.815 E6	-0.6	96	0.00
8 T	DICHLORPROP	1.973	2.057 E6	-4.3	101	0.00
9 T	2,4-D	2.243	2.437 E6	-8.6	105	0.00
10 T	Pentachlorophenol	28.111	28.522 E6	-1.5	98	0.00
11 T	2,4,5-TP (SILVEX)	10.972	11.346 E6	-3.4	102	0.00
12 T	2,4,5-T	10.875	11.146 E6	-2.5	104	0.00
13 T	2,4-DB	2.325	2.374 E6	-2.1	103	0.00
14 T	DINOSEB	8.189	8.478 E6	-3.5	105	0.00
15 T	Picloram	16.393	16.902 E6	-3.1	104	0.00
16 T	DCPA	16.190	14.592 E6	9.9	92	0.00

Signal #2

1 T	Dalapon	615.395	606.860 E3	1.4	101	0.00
2 T	3,5-DICHLORO BENZOIC ACID	688.541	690.527 E3	-0.3	102	0.00
3 T	4-Nitrophenol	282.863	282.082 E3	0.3	102	0.00
4 S	2,4-DCAA	462.153	457.822 E3	0.9	102	0.00
5 T	DICAMBA	1.910	1.949 E6	-2.0	103	0.00
6 T	MCP P	1.532	1.535 E6	-0.2	103	0.00
7 T	MCPA	2.326	2.264 E6	2.7	101	0.00
8 T	DICHLORPROP	475.910	486.765 E3	-2.3	106	0.00
9 T	2,4-D	570.994	610.569 E3	-6.9	110	0.00
10 T	Pentachlorophenol	6.545	6.509 E6	0.6	101	0.00
11 T	2,4,5-TP (SILVEX)	2.555	2.603 E6	-1.9	104	0.00
12 T	2,4,5-T	2.599	2.599 E6	0.0	102	0.00
13 T	2,4-DB	341.930	338.897 E3	0.9	103	0.00
14 T	DINOSEB	1.860	1.872 E6	-0.6	103	0.00
15 T	Picloram	3.932	3.940 E6	-0.2	102	0.00
16 T	DCPA	3.817	3.315 E6	13.2	89	0.00

Evaluate Continuing Calibration Report - Not Found

Signal #2

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS052721\
Data File : PS015092.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 May 2021 21:44
Operator : DD\AJ
Sample : HSTDICV750
Misc :
ALS Vial : 12 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 28 03:45:00 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052721.M
Quant Title : 8080.M
QLast Update : Fri May 28 03:42:14 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0