

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS052920\
 Data File : PS010522.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 May 2020 13:58
 Operator : DD\AJ
 Sample : HSTDICC750
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 29 15:40:56 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
 Quant Title : 8080.M
 QLast Update : Fri May 29 15:39:53 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.230	6.921	582.0E6	398.9E6	750.000	750.000
Target Compounds							
1) T	Dalapon	3.041	2.623	660.1E6	488.3E6	682.500	682.500
2) T	3,5-DICHL...	6.442	6.070	753.7E6	563.5E6	697.500	697.500
3) T	4-Nitroph...	7.060	6.526	248.2E6	221.9E6	682.500	682.500
5) T	DICAMBA	7.423	7.090	2211.1E6	1573.3E6	705.000	705.000
6) T	MCPD	7.593	7.188	172.6E6	129.2E6	70.500	70.500
7) T	MCPA	7.742	7.404	240.8E6	186.2E6	69.750	69.750
8) T	DICHLORPROP	8.115	7.716	526.6E6	397.3E6	705.000	705.000
9) T	2,4-D	8.344	8.011	551.5E6	512.3E6	705.000	705.000
10) T	Pentachlo...	8.657	8.442	7530.3E6	4992.6E6	712.500	712.500
11) T	2,4,5-TP ...	9.207	8.802	3030.8E6	2119.3E6	712.500	712.500
12) T	2,4,5-T	9.491	9.181	2890.3E6	2121.9E6	712.500	712.500
13) T	2,4-DB	10.044	9.687	391.8E6	285.5E6	712.500	712.500
14) T	DINOSEB	11.207	10.022	2292.1E6	1502.3E6	705.000	705.000
15) T	Picloram	11.024	11.002	3708.2E6	2986.2E6	712.500	712.500
16) T	DCPA	11.471	10.951	3847.5E6	2631.1E6	720.000	720.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS052920\
Data File : PS010522.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 May 2020 13:58
Operator : DD\AJ
Sample : HSTDICC750
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 29 15:40:56 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
Quant Title : 8080.M
QLast Update : Fri May 29 15:39:53 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

