

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS053124\
 Data File : PS026894.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 May 2024 18:51
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 06/03/2024
 Supervised By : Ankita Jodhani 06/03/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 01 00:49:22 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS051424.M
 Quant Title : 8080.M
 QLast Update : Wed May 15 02:07:16 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.246	7.772	1823.2E6	1300.3E6	675.666	712.464
Target Compounds							
1) T	Dalapon	2.649	2.713	2443.7E6	2114.2E6	623.929m	625.945
2) T	3,5-DICHL...	6.416	6.720	2414.7E6	1815.1E6	648.246	669.788
3) T	4-Nitroph...	7.048	7.298	1135.4E6	854.0E6	631.034	637.818
5) T	DICAMBA	7.433	7.972	7179.4E6	5544.5E6	657.984	684.497
6) T	MCPD	7.613	8.073	528.5E6	372.9E6	65.908	68.840
7) T	MCPA	7.762	8.318	746.2E6	527.5E6	65.073	67.860
8) T	DICHLORPROP	8.134	8.687	2025.2E6	1407.7E6	635.717	673.678
9) T	2,4-D	8.364	9.018	2277.8E6	1526.2E6	639.284	671.765
10) T	Pentachlo...	8.661	9.548	29584.0E6	20829.1E6	682.481	691.658
11) T	2,4,5-TP ...	9.237	9.922	11422.5E6	8236.9E6	653.847	691.210
12) T	2,4,5-T	9.528	10.343	11968.2E6	8223.3E6	654.110	684.748
13) T	2,4-DB	10.101	10.908	1706.6E6	1038.1E6	665.655m	673.990
14) T	DINOSEB	11.305	11.287	7882.7E6	5609.2E6	662.109m	681.694
15) T	Picloram	11.117	12.385	15260.0E6	11931.1E6	646.119	669.829
16) T	DCPA	11.599	12.331	13634.7E6	9961.2E6	670.193	689.485

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS053124\
Data File : PS026894.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 May 2024 18:51
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :

ECD_S

ClientSampleId :

HSTDCCC750

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 06/03/2024

Supervised By :Ankita Jodhani 06/03/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 01 00:49:22 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS051424.M
Quant Title : 8080.M
QLast Update : Wed May 15 02:07:16 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

