

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060120\
 Data File : PS010560.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Jun 2020 16:10
 Operator : DD\AJ
 Sample : L2773-09
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 01 18:53:04 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
 Quant Title : 8080.M
 QLast Update : Fri May 29 15:43:50 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml



System Monitoring Compounds							
4) S	2,4-DCAA	7.231	6.921	194.0E6	138.5E6	245.291	253.898
Target Compounds							
2) T	3,5-DICHL...	0.000	6.100	0	71765435	N.D.	87.557
3) T	4-Nitroph...	7.009	6.557	2663878	12014969	7.186	36.270 #
5) T	DICAMBA	0.000	7.128	0	7883061	N.D.	3.502
6) T	MCP	7.598	7.185	10159010	12023539	4.112	6.452 #
7) T	MCPA	7.724	7.386	23022534	63679826	6.585	23.431 #
8) T	DICHLORPROP	0.000	7.738	0	23477619	N.D.	40.715
9) T	2,4-D	8.276	7.997	11566805	13334037	14.674	17.585
10) T	Pentachlo...	8.655	8.438	-21903	13361724	N.D.	1.881
11) T	2,4,5-TP ...	9.208	8.843	-2017140	1860350	N.D.	<MDL
12) T	2,4,5-T	9.433	9.172	67541606	45240402	16.607	15.076
13) T	2,4-DB	10.023	9.708	22266313	82509565	40.270	203.114 #
14) T	DINOSEB	11.229	10.036	123.2E6	87246900	38.145	40.553
15) T	Picloram	11.019	11.001	144.5E6	120.5E6	28.305	29.197
16) T	DCPA	11.468	10.948	63705235	18248269	11.890	4.921 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060120\
Data File : PS010560.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Jun 2020 16:10
Operator : DD\AJ
Sample : L2773-09
Misc :
ALS Vial : 14 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 01 18:53:04 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
Quant Title : 8080.M
QLast Update : Fri May 29 15:43:50 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon

R.T.: 2.996 min
Delta R.T.: -0.045 min
Response: 28076058
Conc: 28.56 ng/ml




#1 Dalapon

R.T.: 2.584 min
Delta R.T.: -0.040 min
Response: -33045669
Conc: N.D.




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 6.443 min
Response: 0
Conc: N.D.




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 6.100 min
Delta R.T.: 0.030 min
Response: 71765435
Conc: 87.56 ng/ml




#3 4-Nitrophenol

R.T.: 7.009 min
Delta R.T.: -0.051 min
Response: 2663878
Conc: 7.19 ng/ml




#3 4-Nitrophenol

R.T.: 6.557 min
Delta R.T.: 0.032 min
Response: 12014969
Conc: 36.27 ng/ml




#4 2,4-DCAA

R.T.: 7.231 min
Delta R.T.: 0.000 min
Response: 193991646
Conc: 245.29 ng/ml




#4 2,4-DCAA

R.T.: 6.921 min
Delta R.T.: 0.000 min
Response: 138479940
Conc: 253.90 ng/ml




#5 DICAMBA

R.T.: 0.000 min
Exp R.T.: 7.423 min
Response: 0
Conc: N.D.




#5 DICAMBA

R.T.: 7.128 min
Delta R.T.: 0.037 min
Response: 7883061
Conc: 3.50 ng/ml




#6 MCP

R.T.: 7.598 min
Delta R.T.: -0.002 min
Response: 10159010
Conc: 4.11 ug/ml




#6 MCP

R.T.: 7.185 min
Delta R.T.: -0.012 min
Response: 12023539
Conc: 6.45 ug/ml




#7 MCPA

R.T.: 7.724 min
Delta R.T.: -0.027 min
Response: 23022534
Conc: 6.59 ug/ml




#7 MCPA

R.T.: 7.386 min
Delta R.T.: -0.028 min
Response: 63679826
Conc: 23.43 ug/ml




#8 DICHLORPROP

R.T.: 0.000 min
Exp R.T. : 8.116 min
Response: 0
Conc: N.D.




#8 DICHLORPROP

R.T.: 7.738 min
Delta R.T.: 0.021 min
Response: 23477619
Conc: 40.72 ng/ml




#9 2,4-D

R.T.: 8.276 min
Delta R.T.: -0.069 min
Response: 11566805
Conc: 14.67 ng/ml




#9 2,4-D

R.T.: 7.997 min
Delta R.T.: -0.014 min
Response: 13334037
Conc: 17.59 ng/ml




#10 Pentachlorophenol

R.T.: 8.655 min
Delta R.T.: -0.003 min
Response: -21903
Conc: N.D.




#10 Pentachlorophenol

R.T.: 8.438 min
Delta R.T.: -0.004 min
Response: 13361724
Conc: 1.88 ng/ml




#12 2,4,5-T

R.T.: 9.433 min
Delta R.T.: -0.059 min
Response: 67541606
Conc: 16.61 ng/ml




#12 2,4,5-T

R.T.: 9.172 min
Delta R.T.: -0.010 min
Response: 45240402
Conc: 15.08 ng/ml




#13 2,4-DB

R.T.: 10.023 min
Delta R.T.: -0.022 min
Response: 22266313
Conc: 40.27 ng/ml




#13 2,4-DB

R.T.: 9.708 min
Delta R.T.: 0.020 min
Response: 82509565
Conc: 203.11 ng/ml




#14 DINOSEB

R.T.: 11.229 min
Delta R.T.: 0.021 min
Response: 123249285
Conc: 38.14 ng/ml




#14 DINOSEB

R.T.: 10.036 min
Delta R.T.: 0.014 min
Response: 87246900
Conc: 40.55 ng/ml




#15 Picloram

R.T.: 11.019 min
Delta R.T.: -0.005 min
Response: 144501916
Conc: 28.31 ng/ml




#15 Picloram

R.T.: 11.001 min
Delta R.T.: -0.002 min
Response: 120463772
Conc: 29.20 ng/ml



#16 DCPA

R.T.: 11.468 min
Delta R.T.: -0.004 min
Response: 63705235
Conc: 11.89 ng/ml



#16 DCPA

R.T.: 10.948 min
Delta R.T.: -0.004 min
Response: 18248269
Conc: 4.92 ng/ml