

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060120\
 Data File : PS010564.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Jun 2020 17:42
 Operator : DD\AJ
 Sample : L2773-13
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 01 18:54:07 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
 Quant Title : 8080.M
 QLast Update : Fri May 29 15:43:50 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml



System Monitoring Compounds							
4) S	2,4-DCAA	7.231	6.921	120.7E6	82434989	152.589	151.142
Target Compounds							
2) T	3,5-DICHL...	0.000	6.101	0	81443779	N.D.	99.365
3) T	4-Nitroph...	0.000	6.554	0	11493626	N.D.	34.696
5) T	DICAMBA	0.000	7.053	0	18383129	N.D.	8.167
6) T	MCP	7.546	7.129	9822213	7702616	3.976	4.133
7) T	MCPA	7.723	7.386	15474309	62537245	4.426	23.011 #
8) T	DICHLORPROP	0.000	7.739	0	53787704	N.D.	93.280
9) T	2,4-D	0.000	7.994	0	15854043	N.D.	20.909
10) T	Pentachlo...	0.000	8.410	0	12052912	N.D.	1.697
12) T	2,4,5-T	9.435	9.172	84368473	29732420	20.745	9.908 #
13) T	2,4-DB	10.107	9.706	102.0E6	57703594	184.409	142.049
14) T	DINOSEB	11.174	10.036	122.0E6	78044810	37.745	36.276
15) T	Picloram	10.996	11.019	187.7E6	21298785	36.758	5.162 #
16) T	DCPA	11.482	10.944	57944408	31160929	10.815	8.403


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060120\
Data File : PS010564.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Jun 2020 17:42
Operator : DD\AJ
Sample : L2773-13
Misc :
ALS Vial : 18 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 01 18:54:07 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
Quant Title : 8080.M
QLast Update : Fri May 29 15:43:50 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon

R.T.: 2.998 min
Delta R.T.: -0.043 min
Response: 26622907
Conc: 27.08 ng/ml




#1 Dalapon

R.T.: 2.580 min
Delta R.T.: -0.044 min
Response: -60498168
Conc: N.D.




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 6.443 min
Response: 0
Conc: N.D.




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 6.101 min
Delta R.T.: 0.031 min
Response: 81443779
Conc: 99.36 ng/ml




#3 4-Nitrophenol

R.T.: 0.000 min
Exp R.T.: 7.060 min
Response: 0
Conc: N.D.




#3 4-Nitrophenol

R.T.: 6.554 min
Delta R.T.: 0.029 min
Response: 11493626
Conc: 34.70 ng/ml




#4 2,4-DCAA

R.T.: 7.231 min
Delta R.T.: 0.000 min
Response: 120676657
Conc: 152.59 ng/ml




#4 2,4-DCAA

R.T.: 6.921 min
Delta R.T.: 0.000 min
Response: 82434989
Conc: 151.14 ng/ml




#5 DICAMBA

R.T.: 0.000 min
Exp R.T.: 7.423 min
Response: 0
Conc: N.D.




#5 DICAMBA

R.T.: 7.053 min
Delta R.T.: -0.038 min
Response: 18383129
Conc: 8.17 ng/ml




#6 MCP

R.T.: 7.546 min
Delta R.T.: -0.054 min
Response: 9822213
Conc: 3.98 ug/ml




#6 MCP

R.T.: 7.129 min
Delta R.T.: -0.067 min
Response: 7702616
Conc: 4.13 ug/ml




#7 MCPA

R.T.: 7.723 min
Delta R.T.: -0.028 min
Response: 15474309
Conc: 4.43 ug/ml




#7 MCPA

R.T.: 7.386 min
Delta R.T.: -0.028 min
Response: 62537245
Conc: 23.01 ug/ml




#8 DICHLORPROP

R.T.: 0.000 min
Exp R.T. : 8.116 min
Response: 0
Conc: N.D.




#8 DICHLORPROP

R.T.: 7.739 min
Delta R.T.: 0.022 min
Response: 53787704
Conc: 93.28 ng/ml




#9 2,4-D

R.T.: 0.000 min
Exp R.T. : 8.345 min
Response: 0
Conc: N.D.




#9 2,4-D

R.T.: 7.994 min
Delta R.T.: -0.016 min
Response: 15854043
Conc: 20.91 ng/ml




#10 Pentachlorophenol

R.T.: 0.000 min
Exp R.T. : 8.658 min
Response: 0
Conc: N.D.




#10 Pentachlorophenol

R.T.: 8.410 min
Delta R.T.: -0.032 min
Response: 12052912
Conc: 1.70 ng/ml




#12 2,4,5-T

R.T.: 9.435 min
Delta R.T.: -0.057 min
Response: 84368473
Conc: 20.74 ng/ml




#12 2,4,5-T

R.T.: 9.172 min
Delta R.T.: -0.009 min
Response: 29732420
Conc: 9.91 ng/ml




#13 2,4-DB

R.T.: 10.107 min
Delta R.T.: 0.063 min
Response: 101964691
Conc: 184.41 ng/ml




#13 2,4-DB

R.T.: 9.706 min
Delta R.T.: 0.019 min
Response: 57703594
Conc: 142.05 ng/ml




#14 DINOSEB

R.T.: 11.174 min
Delta R.T.: -0.034 min
Response: 121957735
Conc: 37.75 ng/ml




#14 DINOSEB

R.T.: 10.036 min
Delta R.T.: 0.014 min
Response: 78044810
Conc: 36.28 ng/ml




#15 Picloram

R.T.: 10.996 min
Delta R.T.: -0.028 min
Response: 187656388
Conc: 36.76 ng/ml




#15 Picloram

R.T.: 11.019 min
Delta R.T.: 0.017 min
Response: 21298785
Conc: 5.16 ng/ml



#16 DCPA

R.T.: 11.482 min
Delta R.T.: 0.010 min
Response: 57944408
Conc: 10.81 ng/ml



#16 DCPA

R.T.: 10.944 min
Delta R.T.: -0.008 min
Response: 31160929
Conc: 8.40 ng/ml