

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060120\
 Data File : PS010565.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Jun 2020 18:05
 Operator : DD\AJ
 Sample : L2773-14
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 01 18:54:22 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
 Quant Title : 8080.M
 QLast Update : Fri May 29 15:43:50 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml



System Monitoring Compounds							
4) S	2,4-DCAA	7.230	6.920	141.8E6	88707068	179.283	162.641
Target Compounds							
2) T	3,5-DICHL...	0.000	6.101	0	50880493	N.D.	62.076
3) T	4-Nitroph...	0.000	6.553	0	19926032	N.D.	60.152
5) T	DICAMBA	7.457	7.052	7372531	13548490	2.345	6.019 #
6) T	MCPD	7.593	0.000	10559421	0	4.274	N.D. #
7) T	MCPA	7.742	7.387	24936897	47564183	7.133	17.501 #
8) T	DICHLORPROP	8.130	7.719	12254601	20679087	16.169	35.862 #
9) T	2,4-D	0.000	7.992	0	16272489	N.D.	21.460
10) T	Pentachlo...	8.691	8.411	29746943	10261285	2.789	1.444 #
12) T	2,4,5-T	9.435	9.172	69002296	27656840	16.966	9.216 #
13) T	2,4-DB	10.106	9.707	82727114	60613274	149.617	149.212
14) T	DINOSEB	11.231	10.036	87653444	85370067	27.128	39.680 #
15) T	Picloram	10.996	11.015	181.4E6	10518328	35.525	2.549 #
16) T	DCPA	11.493	10.946	40814014	22728871	7.617	6.129


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060120\
Data File : PS010565.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Jun 2020 18:05
Operator : DD\AJ
Sample : L2773-14
Misc :
ALS Vial : 19 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 01 18:54:22 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
Quant Title : 8080.M
QLast Update : Fri May 29 15:43:50 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T.: 6.443 min
Response: 0
Conc: N.D.




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 6.101 min
Delta R.T.: 0.030 min
Response: 50880493
Conc: 62.08 ng/ml




#3 4-Nitrophenol

R.T.: 0.000 min
Exp R.T.: 7.060 min
Response: 0
Conc: N.D.




#3 4-Nitrophenol

R.T.: 6.553 min
Delta R.T.: 0.028 min
Response: 19926032
Conc: 60.15 ng/ml




#4 2,4-DCAA

R.T.: 7.230 min
Delta R.T.: 0.000 min
Response: 141787904
Conc: 179.28 ng/ml




#4 2,4-DCAA

R.T.: 6.920 min
Delta R.T.: -0.001 min
Response: 88707068
Conc: 162.64 ng/ml




#5 DICAMBA

R.T.: 7.457 min
Delta R.T.: 0.033 min
Response: 7372531
Conc: 2.34 ng/ml




#5 DICAMBA

R.T.: 7.052 min
Delta R.T.: -0.039 min
Response: 13548490
Conc: 6.02 ng/ml




#6 MCPP

R.T.: 7.593 min
Delta R.T.: -0.007 min
Response: 10559421
Conc: 4.27 ug/ml




#6 MCPP

R.T.: 0.000 min
Exp R.T.: 7.196 min
Response: 0
Conc: N.D.




#7 MCPA

R.T.: 7.742 min
Delta R.T.: -0.009 min
Response: 24936897
Conc: 7.13 ug/ml




#7 MCPA

R.T.: 7.387 min
Delta R.T.: -0.027 min
Response: 47564183
Conc: 17.50 ug/ml




#8 DICHLORPROP

R.T.: 8.130 min
Delta R.T.: 0.014 min
Response: 12254601
Conc: 16.17 ng/ml




#8 DICHLORPROP

R.T.: 7.719 min
Delta R.T.: 0.001 min
Response: 20679087
Conc: 35.86 ng/ml




#9 2,4-D

R.T.: 0.000 min
Exp R.T.: 8.345 min
Response: 0
Conc: N.D.




#9 2,4-D

R.T.: 7.992 min
Delta R.T.: -0.019 min
Response: 16272489
Conc: 21.46 ng/ml




#10 Pentachlorophenol

R.T.: 8.691 min
Delta R.T.: 0.034 min
Response: 29746943
Conc: 2.79 ng/ml




#10 Pentachlorophenol

R.T.: 8.411 min
Delta R.T.: -0.032 min
Response: 10261285
Conc: 1.44 ng/ml




#11 2,4,5-TP (SILVEX)

R.T.: 9.245 min
Delta R.T.: 0.037 min
Response: 7848865
Conc: 1.84 ng/ml




#11 2,4,5-TP (SILVEX)

R.T.: 8.842 min
Delta R.T.: 0.040 min
Response: -58395579
Conc: N.D.




#12 2,4,5-T

R.T.: 9.435 min
Delta R.T.: -0.057 min
Response: 69002296
Conc: 16.97 ng/ml




#12 2,4,5-T

R.T.: 9.172 min
Delta R.T.: -0.010 min
Response: 27656840
Conc: 9.22 ng/ml




#13 2,4-DB

R.T.: 10.106 min
Delta R.T.: 0.062 min
Response: 82727114
Conc: 149.62 ng/ml




#13 2,4-DB

R.T.: 9.707 min
Delta R.T.: 0.019 min
Response: 60613274
Conc: 149.21 ng/ml




#14 DINOSEB

R.T.: 11.231 min
Delta R.T.: 0.024 min
Response: 87653444
Conc: 27.13 ng/ml




#14 DINOSEB

R.T.: 10.036 min
Delta R.T.: 0.014 min
Response: 85370067
Conc: 39.68 ng/ml




#15 Picloram

R.T.: 10.996 min
Delta R.T.: -0.028 min
Response: 181361719
Conc: 35.53 ng/ml




#15 Picloram

R.T.: 11.015 min
Delta R.T.: 0.013 min
Response: 10518328
Conc: 2.55 ng/ml



#16 DCPA

R.T.: 11.493 min
Delta R.T.: 0.021 min
Response: 40814014
Conc: 7.62 ng/ml



#16 DCPA

R.T.: 10.946 min
Delta R.T.: -0.006 min
Response: 22728871
Conc: 6.13 ng/ml