

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060120\
 Data File : PS010585.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jun 2020 01:43
 Operator : DD\AJ
 Sample : PB129214BS
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB129214BS

Manual Integrations
 APPROVED

Sohil
 6/2/2020 11:30:15 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 02 06:48:26 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
 Quant Title : 8080.M
 QLast Update : Fri May 29 15:43:50 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.229	6.920	416.7E6	295.0E6	526.866	540.796m
Target Compounds							
1) T	Dalapon	3.036	2.622	431.0E6	394.8E6	438.367	539.172
2) T	3,5-DICHL...	6.441	6.069	577.8E6	412.3E6	528.526	502.973m
3) T	4-Nitroph...	7.059	6.525	187.6E6	144.1E6	505.991	434.927
5) T	DICAMBA	7.422	7.090	1503.0E6	1077.9E6	477.980	478.884
6) T	MCPD	7.590	7.186	122.9E6	90434703	49.740	48.530
7) T	MCPA	7.738	7.399	174.9E6	159.7E6	50.017	58.767
8) T	DICHLORPROP	8.114	7.716	390.2E6	289.1E6	514.796	501.406
9) T	2,4-D	8.343	8.009	449.8E6	389.8E6	570.578	514.010
10) T	Pentachlo...	8.656	8.441	6022.9E6	3826.7E6	564.681	538.641
11) T	2,4,5-TP ...	9.205	8.801	2370.4E6	1598.6E6	554.238	532.280m
12) T	2,4,5-T	9.490	9.180	2130.3E6	1539.5E6	523.811	513.010
13) T	2,4-DB	10.043	9.686	280.8E6	204.1E6	507.828	502.458
14) T	DINOSEB	11.206	10.021	1654.3E6	1128.8E6	511.985	524.668
15) T	Picloram	11.022	11.000	1894.0E6	1444.8E6	371.002	350.166
16) T	DCPA	11.471	10.949	2651.4E6	1953.9E6	494.854	526.908

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060120\
Data File : PS010585.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jun 2020 01:43
Operator : DD\AJ
Sample : PB129214BS
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB129214BS

Manual Integrations
APPROVED

Sohil
6/2/2020 11:30:15 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 02 06:48:26 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
Quant Title : 8080.M
QLast Update : Fri May 29 15:43:50 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

