

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060120\
 Data File : PS010595.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jun 2020 05:33
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 02 06:51:50 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
 Quant Title : 8080.M
 QLast Update : Fri May 29 15:43:50 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.230	6.921	632.7E6	427.3E6	800.023	783.527
Target Compounds							
1) T	Dalapon	3.042	2.624	661.4E6	479.0E6	672.670	654.293
2) T	3,5-DICHL...	6.443	6.070	823.1E6	607.3E6	752.917	740.922
3) T	4-Nitroph...	7.059	6.525	281.4E6	238.8E6	759.062	720.995
5) T	DICAMBA	7.422	7.090	2297.0E6	1619.5E6	730.498	719.504
6) T	MCPD	7.593	7.189	180.8E6	137.0E6	73.173	73.494
7) T	MCPA	7.743	7.405	252.4E6	197.6E6	72.184	72.705
8) T	DICHLORPROP	8.114	7.716	564.8E6	422.1E6	745.204	731.981
9) T	2,4-D	8.344	8.010	601.7E6	525.3E6	763.337	692.831
10) T	Pentachlo...	8.656	8.441	8164.4E6	5036.7E6	765.451	708.947
11) T	2,4,5-TP ...	9.206	8.802	3252.2E6	2215.9E6	760.417	737.831
12) T	2,4,5-T	9.490	9.180	3126.4E6	2230.9E6	768.724	743.405
13) T	2,4-DB	10.042	9.687	431.9E6	299.7E6	781.158	737.718
14) T	DINOSEB	11.206	10.021	2383.8E6	1590.2E6	737.772	739.118
15) T	Picloram	11.023	11.002	3524.5E6	2696.3E6	690.386	653.502
16) T	DCPA	11.471	10.950	4095.1E6	2786.6E6	764.308	751.464

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060120\
Data File : PS010595.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jun 2020 05:33
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 02 06:51:50 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
Quant Title : 8080.M
QLast Update : Fri May 29 15:43:50 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

