

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060221\
 Data File : PS015161.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jun 2021 11:59
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 02 17:47:39 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052821.M
 Quant Title : 8080.M
 QLast Update : Fri May 28 14:34:22 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.224	6.783	1757.9E6	385.6E6	697.457	719.709
Target Compounds							
1) T	Dalapon	3.036	2.520	1784.1E6	435.2E6	678.790	664.995
2) T	3,5-DICHL...	6.435	5.969	2209.9E6	539.3E6	640.851	667.598
3) T	4-Nitroph...	7.049	6.399	901.8E6	210.1E6	639.739	660.301
5) T	DICAMBA	7.417	6.948	6381.6E6	1527.4E6	669.510	690.217
6) T	MCPD	7.588	7.046	503.6E6	123.7E6	69.983	70.005
7) T	MCPA	7.737	7.259	690.9E6	178.6E6	68.838	68.726
8) T	DICHLORPROP	8.112	7.570	1532.6E6	372.4E6	665.432	685.422
9) T	2,4-D	8.339	7.861	1701.1E6	441.3E6	667.607	685.232
10) T	Pentachlo...	8.653	8.292	21686.6E6	5081.8E6	669.943	690.349
11) T	2,4,5-TP ...	9.204	8.654	8424.7E6	2039.4E6	678.806	702.769
12) T	2,4,5-T	9.486	9.031	8267.9E6	2061.0E6	688.365	711.910
13) T	2,4-DB	10.038	9.538	1768.0E6	271.2E6	694.266	703.152
14) T	DINOSEB	11.203	9.867	5980.3E6	1464.3E6	748.509	801.855
15) T	Picloram	11.016	10.845	11715.4E6	3029.3E6	707.164	729.455
16) T	DCPA	11.471	10.801	11835.6E6	3104.6E6	701.908	720.159

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060221\
Data File : PS015161.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jun 2021 11:59
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 02 17:47:39 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052821.M
Quant Title : 8080.M
QLast Update : Fri May 28 14:34:22 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

