

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060321\
 Data File : PS015174.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Jun 2021 13:28
 Operator : DD\AJ
 Sample : M2569-02MSD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 C0AA1MSD

Manual Integrations
 APPROVED

Ankita
 6/4/2021 8:40:44 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 03 15:43:47 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052821.M
 Quant Title : 8080.M
 QLast Update : Fri May 28 14:34:22 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.227	6.786	1218.8E6	265.3E6	483.570	495.232
Target Compounds							
1) T	Dalapon	3.035	2.521	1046.2E6	391.5E6	398.036	598.218m#
2) T	3,5-DICHL...	6.437	5.970	1648.8E6	387.2E6	478.146	479.306m
3) T	4-Nitroph...	7.053	6.402	573.0E6	139.6E6	406.466	438.776
5) T	DICAMBA	7.420	6.950	4329.0E6	1035.6E6	454.169	467.954
6) T	MCPD	7.589	7.046	373.6E6	83086723	51.909	47.018
7) T	MCPA	7.737	7.258	490.2E6	126.0E6	48.847	48.493
8) T	DICHLORPROP	8.115	7.573	1100.3E6	250.4E6	477.729	460.947
9) T	2,4-D	8.342	7.863	1344.4E6	353.9E6	527.605	549.568m
10) T	Pentachlo...	8.656	8.295	16151.6E6	3798.0E6	498.957	515.947
11) T	2,4,5-TP ...	9.207	8.657	6457.0E6	1540.3E6	520.264	530.792
12) T	2,4,5-T	9.490	9.034	5759.1E6	1458.3E6	479.493	503.721
13) T	2,4-DB	10.044	9.541	793.2E6	188.6E6	311.498	489.010 #
14) T	DINOSEB	11.206	9.870	4260.9E6	1011.9E6	533.307	554.119
15) T	Picloram	11.019	10.848	6887.7E6	1774.2E6	415.756	427.212
16) T	DCPA	11.474	10.804	7046.4E6	1892.6E6	417.886	439.025

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060321\
Data File : PS015174.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jun 2021 13:28
Operator : DD\AJ
Sample : M2569-02MSD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
C0AA1MSD

Manual Integrations
APPROVED

Ankita
6/4/2021 8:40:44 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 03 15:43:47 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052821.M
Quant Title : 8080.M
QLast Update : Fri May 28 14:34:22 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

