

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060322\
 Data File : PS019595.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jun 2022 14:08
 Operator : AJ\MA
 Sample : PB145272BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB145272BS

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 06/15/2022
 Supervised By : Ankita Jodhani 06/15/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 02 16:19:01 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052422.M
 Quant Title : 8080.M
 QLast Update : Wed May 25 02:05:34 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.198	6.782	1381.4E6	480.9E6	575.795	620.619
Target Compounds							
1) T	Dalapon	3.015	2.532	2208.4E6	627.0E6	435.694	624.886 #
2) T	3,5-DICHL...	6.412	5.962	1924.8E6	805.8E6	608.982	672.273
3) T	4-Nitroph...	7.029	6.400	573.0E6	211.2E6	583.874	493.438
5) T	DICAMBA	7.388	6.947	5293.7E6	1989.7E6	557.360	582.054
6) T	MCP	7.557	7.043	490.8E6	163.7E6	65.394	60.462
7) T	MCPA	7.705	7.254	582.2E6	235.7E6	53.884	61.329
8) T	DICHLORPROP	8.079	7.566	1410.5E6	546.1E6	550.837	618.153
9) T	2,4-D	8.308	7.857	1730.1E6	648.9E6	605.032	608.374m
10) T	Pentachlo...	8.615	8.283	21248.2E6	7428.9E6	614.308	637.330
11) T	2,4,5-TP ...	9.165	8.643	9259.8E6	3133.1E6	589.761	623.232
12) T	2,4,5-T	9.451	9.021	10001.2E6	2816.0E6	531.421	598.745
13) T	2,4-DB	10.003	9.527	1651.7E6	433.6E6	408.222	701.584 #
14) T	DINOSEB	11.156	9.851	8625.1E6	1730.4E6	557.640	487.088
15) T	Picloram	10.980	10.832	15503.6E6	3185.2E6	522.405	495.193
16) T	DCPA	11.426	10.780	12753.7E6	3880.2E6	529.109	628.633

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060322\
Data File : PS019595.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jun 2022 14:08
Operator : AJ\MA
Sample : PB145272BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB145272BS

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 06/15/2022
Supervised By : Ankita Jodhani 06/15/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 02 16:19:01 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052422.M
Quant Title : 8080.M
QLast Update : Wed May 25 02:05:34 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

