

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060420\
 Data File : PS010653.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Jun 2020 12:53
 Operator : DD\AJ
 Sample : L2839-06
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 04 13:24:26 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
 Quant Title : 8080.M
 QLast Update : Fri May 29 15:43:50 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.229	6.921	229.3E6	237.0E6	289.969	434.546 #
Target Compounds							
2) T	3,5-DICHL...	0.000	6.101	0	89373302	N.D.	109.039
3) T	4-Nitroph...	7.065	0.000	12940633	0	34.906	N.D. #
5) T	DICAMBA	0.000	7.054	0	11515378	N.D.	5.116
6) T	MCP	7.594	0.000	13200069	0	5.343	N.D. #
7) T	MCPA	7.779	7.415	36208322	19488012	10.357	7.171 #
8) T	DICHLORPROP	8.132	7.739	624855	23528823	<MDL	40.804 #
10) T	Pentachlo...	8.655	8.441	93726594	112.8E6	8.787	15.876 #
11) T	2,4,5-TP ...	9.241	8.843	7275344	687672	1.701	<MDL #
12) T	2,4,5-T	9.516	9.171	4402984	16029658	1.083	5.342 #
13) T	2,4-DB	0.000	9.707	0	37753376	N.D.	92.938
14) T	DINOSEB	11.230	10.037	44741318	83330180	13.847	38.732 #
15) T	Picloram	10.996	11.001	157.6E6	24297242	30.869	5.889 #
16) T	DCPA	11.489	10.945	40489998	11602893	7.557	3.129 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060420\
Data File : PS010653.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Jun 2020 12:53
Operator : DD\AJ
Sample : L2839-06
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 04 13:24:26 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
Quant Title : 8080.M
QLast Update : Fri May 29 15:43:50 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.994 min
Delta R.T.: -0.047 min
Response: 30577024
Conc: 31.10 ng/ml

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.580 min
Delta R.T.: -0.043 min
Response: -24863175
Conc: N.D.




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 6.443 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.101 min
Delta R.T.: 0.030 min
Response: 89373302
Conc: 109.04 ng/ml



#3 4-Nitrophenol


R.T.: 7.065 min
Delta R.T.: 0.004 min
Response: 12940633
Conc: 34.91 ng/ml

Instrument :
ECD_S
ClientSampleId :




#3 4-Nitrophenol

R.T.: 0.000 min
Exp R.T. : 6.525 min
Response: 0
Conc: N.D.




#4 2,4-DCAA

R.T.: 7.229 min
Delta R.T.: -0.002 min
Response: 229325292
Conc: 289.97 ng/ml



#4 2,4-DCAA


R.T.: 6.921 min
Delta R.T.: 0.000 min
Response: 237008444
Conc: 434.55 ng/ml



#5 DICAMBA


R.T.: 0.000 min
Exp R.T.: 7.423 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#5 DICAMBA

R.T.: 7.054 min
Delta R.T.: -0.037 min
Response: 11515378
Conc: 5.12 ng/ml




#6 MCP

R.T.: 7.594 min
Delta R.T.: -0.006 min
Response: 13200069
Conc: 5.34 ug/ml



#6 MCP


R.T.: 0.000 min
Exp R.T.: 7.196 min
Response: 0
Conc: N.D.



#7 MCPA


R.T.: 7.779 min
Delta R.T.: 0.028 min
Response: 36208322
Conc: 10.36 ug/ml

Instrument :
ECD_S
ClientSampleId :




#7 MCPA

R.T.: 7.415 min
Delta R.T.: 0.001 min
Response: 19488012
Conc: 7.17 ug/ml




#8 DICHLORPROP

R.T.: 8.132 min
Delta R.T.: 0.016 min
Response: 624855
Conc: N.D.



#8 DICHLORPROP


R.T.: 7.739 min
Delta R.T.: 0.021 min
Response: 23528823
Conc: 40.80 ng/ml



#10 Pentachlorophenol


R.T.: 8.655 min
Delta R.T.: -0.003 min
Response: 93726594
Conc: 8.79 ng/ml

Instrument :
ECD_S
ClientSampleId :




#10 Pentachlorophenol

R.T.: 8.441 min
Delta R.T.: -0.002 min
Response: 112791617
Conc: 15.88 ng/ml




#11 2,4,5-TP (SILVEX)

R.T.: 9.241 min
Delta R.T.: 0.034 min
Response: 7275344
Conc: 1.70 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 8.843 min
Delta R.T.: 0.041 min
Response: 687672
Conc: N.D.



#12 2,4,5-T


R.T.: 9.516 min
Delta R.T.: 0.024 min
Response: 4402984
Conc: 1.08 ng/ml

Instrument :
ECD_S
ClientSampleId :




#12 2,4,5-T

R.T.: 9.171 min
Delta R.T.: -0.010 min
Response: 16029658
Conc: 5.34 ng/ml




#13 2,4-DB

R.T.: 0.000 min
Exp R.T. : 10.044 min
Response: 0
Conc: N.D.



#13 2,4-DB


R.T.: 9.707 min
Delta R.T.: 0.020 min
Response: 37753376
Conc: 92.94 ng/ml



#14 DINOSEB


R.T.: 11.230 min
Delta R.T.: 0.022 min
Response: 44741318
Conc: 13.85 ng/ml

Instrument :
ECD_S
ClientSampleId :




#14 DINOSEB

R.T.: 10.037 min
Delta R.T.: 0.015 min
Response: 83330180
Conc: 38.73 ng/ml




#15 Picloram

R.T.: 10.996 min
Delta R.T.: -0.028 min
Response: 157588418
Conc: 30.87 ng/ml



#15 Picloram


R.T.: 11.001 min
Delta R.T.: -0.001 min
Response: 24297242
Conc: 5.89 ng/ml



#16 DCPA

R.T.: 11.489 min
Delta R.T.: 0.017 min
Response: 40489998
Conc: 7.56 ng/ml

Instrument :
ECD_S
ClientSampleId :



#16 DCPA

R.T.: 10.945 min
Delta R.T.: -0.007 min
Response: 11602893
Conc: 3.13 ng/ml