

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_S\Data\PS060420\  
 Data File : PS010657.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 04 Jun 2020 14:25  
 Operator : DD\AJ  
 Sample : HSTDCCC750  
 Misc :  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 ECD\_S  
 ClientSampleId :  
 HSTDCCC750

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Jun 04 15:14:32 2020  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_S\Method\PS052920.M  
 Quant Title : 8080.M  
 QLast Update : Fri May 29 15:43:50 2020  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 2 µl  
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2  
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

|                             | Compound     | RT#1   | RT#2   | Resp#1   | Resp#2   | ng/ml   | ng/ml   |
|-----------------------------|--------------|--------|--------|----------|----------|---------|---------|
| -----                       |              |        |        |          |          |         |         |
| System Monitoring Compounds |              |        |        |          |          |         |         |
| 4) S                        | 2,4-DCAA     | 7.230  | 6.922  | 628.2E6  | 419.3E6  | 794.284 | 768.778 |
| Target Compounds            |              |        |        |          |          |         |         |
| 1) T                        | Dalapon      | 3.041  | 2.624  | 691.3E6  | 485.3E6  | 703.079 | 662.829 |
| 2) T                        | 3,5-DICHL... | 6.442  | 6.071  | 807.4E6  | 589.1E6  | 738.597 | 718.732 |
| 3) T                        | 4-Nitroph... | 7.060  | 6.526  | 275.7E6  | 236.5E6  | 743.747 | 714.056 |
| 5) T                        | DICAMBA      | 7.422  | 7.090  | 2346.1E6 | 1621.1E6 | 746.106 | 720.215 |
| 6) T                        | MCPD         | 7.592  | 7.189  | 183.1E6  | 135.2E6  | 74.137  | 72.531  |
| 7) T                        | MCPA         | 7.743  | 7.405  | 253.4E6  | 197.3E6  | 72.483  | 72.580  |
| 8) T                        | DICHLORPROP  | 8.114  | 7.717  | 565.5E6  | 413.7E6  | 746.146 | 717.474 |
| 9) T                        | 2,4-D        | 8.344  | 8.010  | 601.5E6  | 511.6E6  | 763.020 | 674.751 |
| 10) T                       | Pentachlo... | 8.656  | 8.443  | 8115.0E6 | 5143.1E6 | 760.822 | 723.928 |
| 11) T                       | 2,4,5-TP ... | 9.206  | 8.802  | 3250.0E6 | 2197.2E6 | 759.918 | 731.589 |
| 12) T                       | 2,4,5-T      | 9.490  | 9.181  | 3152.1E6 | 2217.2E6 | 775.039 | 738.825 |
| 13) T                       | 2,4-DB       | 10.043 | 9.687  | 423.7E6  | 303.4E6  | 766.365 | 746.774 |
| 14) T                       | DINOSEB      | 11.206 | 10.022 | 2423.3E6 | 1614.1E6 | 750.006 | 750.230 |
| 15) T                       | Picloram     | 11.023 | 11.002 | 3744.2E6 | 2796.5E6 | 733.412 | 677.774 |
| 16) T                       | DCPA         | 11.471 | 10.950 | 4033.4E6 | 2751.5E6 | 752.780 | 742.010 |
| -----                       |              |        |        |          |          |         |         |

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_S\Data\PS060420\  
Data File : PS010657.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 04 Jun 2020 14:25  
Operator : DD\AJ  
Sample : HSTDCCC750  
Misc :  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
ECD\_S  
ClientSampleId :  
HSTDCCC750

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Jun 04 15:14:32 2020  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_S\Method\PS052920.M  
Quant Title : 8080.M  
QLast Update : Fri May 29 15:43:50 2020  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 2 µl  
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2  
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

