

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060721\
 Data File : PS015266.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jun 2021 02:01
 Operator : DD\AJ
 Sample : M2589-26
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 08 03:58:17 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052821.M
 Quant Title : 8080.M
 QLast Update : Fri May 28 14:34:22 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.225	6.784	1391.2E6	275.8E6	551.980	514.685
Target Compounds							
1) T	Dalapon	3.005	2.512	97342506	36479849	37.035	55.741 #
2) T	3,5-DICHL...	6.439	5.950	601634	2759712	<MDL	3.416 #
3) T	4-Nitroph...	7.064	6.389	11408121	14969391	8.093	47.048 #
5) T	DICAMBA	7.406	6.996f	14594963	5183525	1.531	2.342 #
6) T	MCPD	7.590	7.035	18727495	7462433	2.602	4.223 #
7) T	MCPA	7.724	7.312f	9863308	5547204	<MDL	2.135 #
8) T	DICHLORPROP	8.125	7.587	15857446	6405664	6.885	11.791 #
9) T	2,4-D	0.000	7.849	0	7716571	N.D.	11.982
10) T	Pentachlo...	8.653	8.285	33393401	12817854	1.032	1.741 #
12) T	2,4,5-T	9.432f	9.094f	136.7E6	10067046	11.377	3.477 #
13) T	2,4-DB	10.009	9.544	45391066	3963766	17.825	10.275 #
14) T	DINOSEB	11.165f	9.908	177.8E6	15004442	22.248	8.216 #
15) T	Picloram	10.991	10.859	185.5E6	11423176	11.197	2.751 #
16) T	DCPA	11.461	10.795	14847156	16680543	<MDL	3.869 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060721\
Data File : PS015266.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jun 2021 02:01
Operator : DD\AJ
Sample : M2589-26
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 08 03:58:17 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052821.M
Quant Title : 8080.M
QLast Update : Fri May 28 14:34:22 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 3.005 min
Delta R.T.: -0.032 min
Response: 97342506
Conc: 37.04 ng/ml

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.512 min
Delta R.T.: -0.008 min
Response: 36479849
Conc: 55.74 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 6.439 min
Delta R.T.: 0.002 min
Response: 601634
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 5.950 min
Delta R.T.: -0.021 min
Response: 2759712
Conc: 3.42 ng/ml



#3 4-Nitrophenol


R.T.: 7.064 min
Delta R.T.: 0.013 min
Response: 11408121
Conc: 8.09 ng/ml

Instrument :
ECD_S
ClientSampleId :




#3 4-Nitrophenol

R.T.: 6.389 min
Delta R.T.: -0.013 min
Response: 14969391
Conc: 47.05 ng/ml




#4 2,4-DCAA

R.T.: 7.225 min
Delta R.T.: 0.000 min
Response: 1391195483
Conc: 551.98 ng/ml



#4 2,4-DCAA


R.T.: 6.784 min
Delta R.T.: -0.002 min
Response: 275753450
Conc: 514.69 ng/ml



#5 DICAMBA


R.T.: 7.406 min
Delta R.T.: -0.014 min
Response: 14594963
Conc: 1.53 ng/ml

Instrument :
ECD_S
ClientSampleId :




#5 DICAMBA

R.T.: 6.996 min
Delta R.T.: 0.044 min
Response: 5183525
Conc: 2.34 ng/ml




#6 MCP

R.T.: 7.590 min
Delta R.T.: 0.000 min
Response: 18727495
Conc: 2.60 ug/ml



#6 MCP


R.T.: 7.035 min
Delta R.T.: -0.014 min
Response: 7462433
Conc: 4.22 ug/ml



#7 MCPA


R.T.: 7.724 min
Delta R.T.: -0.014 min
Response: 9863308
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#7 MCPA

R.T.: 7.312 min
Delta R.T.: 0.050 min
Response: 5547204
Conc: 2.13 ug/ml




#8 DICHLORPROP

R.T.: 8.125 min
Delta R.T.: 0.012 min
Response: 15857446
Conc: 6.89 ng/ml



#8 DICHLORPROP


R.T.: 7.587 min
Delta R.T.: 0.013 min
Response: 6405664
Conc: 11.79 ng/ml



#9 2,4-D


R.T.: 0.000 min
Exp R.T.: 8.341 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#9 2,4-D

R.T.: 7.849 min
Delta R.T.: -0.015 min
Response: 7716571
Conc: 11.98 ng/ml




#10 Pentachlorophenol

R.T.: 8.653 min
Delta R.T.: -0.003 min
Response: 33393401
Conc: 1.03 ng/ml



#10 Pentachlorophenol


R.T.: 8.285 min
Delta R.T.: -0.012 min
Response: 12817854
Conc: 1.74 ng/ml



#12 2,4,5-T


R.T.: 9.432 min
Delta R.T.: -0.058 min
Response: 136651193
Conc: 11.38 ng/ml

Instrument :
ECD_S
ClientSampleId :




#12 2,4,5-T

R.T.: 9.094 min
Delta R.T.: 0.059 min
Response: 10067046
Conc: 3.48 ng/ml




#13 2,4-DB

R.T.: 10.009 min
Delta R.T.: -0.031 min
Response: 45391066
Conc: 17.82 ng/ml



#13 2,4-DB


R.T.: 9.544 min
Delta R.T.: 0.001 min
Response: 3963766
Conc: 10.28 ng/ml



#14 DINOSEB


R.T.: 11.165 min
Delta R.T.: -0.041 min
Response: 177750649
Conc: 22.25 ng/ml

Instrument :
ECD_S
ClientSampleId :




#14 DINOSEB

R.T.: 9.908 min
Delta R.T.: 0.036 min
Response: 15004442
Conc: 8.22 ng/ml




#15 Picloram

R.T.: 10.991 min
Delta R.T.: -0.028 min
Response: 185492713
Conc: 11.20 ng/ml



#15 Picloram


R.T.: 10.859 min
Delta R.T.: 0.008 min
Response: 11423176
Conc: 2.75 ng/ml



#16 DCPA

R.T.: 11.461 min
Delta R.T.: -0.014 min
Response: 14847156
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :



#16 DCPA

R.T.: 10.795 min
Delta R.T.: -0.011 min
Response: 16680543
Conc: 3.87 ng/ml