

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060820\
 Data File : PS010705.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jun 2020 09:23
 Operator : DD\AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 08 11:38:14 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
 Quant Title : 8080.M
 QLast Update : Fri May 29 15:43:50 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds						
9) T	2,4-D	0.000	7.946	0	14577938	N.D. 19.226
15) T	Picloram	11.032	11.010	8662290	10570564	1.697 2.562 #
16) T	DCPA	0.000	11.010	0	10570564	N.D. 2.851



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060820\
Data File : PS010705.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jun 2020 09:23
Operator : DD\AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 08 11:38:14 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
Quant Title : 8080.M
QLast Update : Fri May 29 15:43:50 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA


R.T.: 0.000 min
Exp R.T. : 7.231 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#4 2,4-DCAA

R.T.: 0.000 min
Exp R.T. : 6.921 min
Response: 0
Conc: N.D.




#9 2,4-D

R.T.: 0.000 min
Exp R.T. : 8.345 min
Response: 0
Conc: N.D.



#9 2,4-D


R.T.: 7.946 min
Delta R.T.: -0.064 min
Response: 14577938
Conc: 19.23 ng/ml



#15 Picloram


R.T.: 11.032 min
Delta R.T.: 0.008 min
Response: 8662290
Conc: 1.70 ng/ml

Instrument :
ECD_S
ClientSampleId :




#15 Picloram

R.T.: 11.010 min
Delta R.T.: 0.007 min
Response: 10570564
Conc: 2.56 ng/ml



#16 DCPA

R.T.: 0.000 min
Exp R.T. : 11.472 min
Response: 0
Conc: N.D.



#16 DCPA

R.T.: 11.010 min
Delta R.T.: 0.058 min
Response: 10570564
Conc: 2.85 ng/ml