

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060822\
 Data File : PS019637.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jun 2022 14:28
 Operator : AJ\MA
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 08 18:05:11 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052422.M
 Quant Title : 8080.M
 QLast Update : Wed May 25 02:05:34 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.200	6.784	1014.0E6	405.7E6	422.631	523.610
Target Compounds							
1) T	Dalapon	3.042	0.000	677502	0	<MDL	N.D. #
2) T	3,5-DICHL...	6.368f	0.000	212968	0	<MDL	N.D. #
7) T	MCPA	7.663	0.000	852347	0	<MDL	N.D. #
10) T	Pentachlo...	8.619	8.287	620290	1502130	<MDL	<MDL #
11) T	2,4,5-TP ...	9.123f	8.622	1514020	906994	<MDL	<MDL #
12) T	2,4,5-T	0.000	9.065	0	354474	N.D.	<MDL
13) T	2,4-DB	9.983	9.474f	1687841	1355460	<MDL	2.193 #
14) T	DINOSEB	11.155	0.000	10354346	0	<MDL	N.D. #
15) T	Picloram	10.976	10.784f	810012	1048396	<MDL	<MDL #
16) T	DCPA	11.429	10.784	2078452	1048396	<MDL	<MDL #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060822\
Data File : PS019637.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jun 2022 14:28
Operator : AJ\MA
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 08 18:05:11 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052422.M
Quant Title : 8080.M
QLast Update : Wed May 25 02:05:34 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA


R.T.: 7.200 min
Delta R.T.: -0.001 min
Response: 1013963457
Conc: 422.63 ng/ml

Instrument :
ECD_S
ClientSampleId :




#4 2,4-DCAA

R.T.: 6.784 min
Delta R.T.: 0.000 min
Response: 405729662
Conc: 523.61 ng/ml



#13 2,4-DB

R.T.: 9.983 min
Delta R.T.: -0.027 min
Response: 1687841
Conc: N.D.



#13 2,4-DB

R.T.: 9.474 min
Delta R.T.: -0.057 min
Response: 1355460
Conc: 2.19 ng/ml