

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060822\
 Data File : PS019661.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jun 2022 02:09
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 09 03:45:03 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052422.M
 Quant Title : 8080.M
 QLast Update : Wed May 25 02:05:34 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.198	6.782	1767.7E6	598.0E6	736.810	771.702
Target Compounds							
1) T	Dalapon	3.014	2.525	3202.6E6	698.9E6	631.831	696.544
2) T	3,5-DICHL...	6.412	5.962	2182.0E6	876.6E6	690.375	731.267
3) T	4-Nitroph...	7.029	6.400	742.6E6	305.6E6	756.786	713.853
5) T	DICAMBA	7.387	6.946	6571.3E6	2533.2E6	691.875	741.027
6) T	MCPD	7.556	7.043	523.9E6	200.1E6	69.799	73.931
7) T	MCPA	7.705	7.256	738.8E6	277.9E6	68.381	72.332
8) T	DICHLORPROP	8.078	7.565	1736.0E6	646.1E6	677.983	731.422
9) T	2,4-D	8.307	7.858	2014.5E6	776.1E6	704.501	727.662
10) T	Pentachlo...	8.614	8.282	23957.4E6	8650.5E6	692.633	742.131
11) T	2,4,5-TP ...	9.164	8.643	11062.1E6	3737.9E6	704.552	743.550
12) T	2,4,5-T	9.449	9.021	13100.6E6	3502.2E6	696.111	744.641
13) T	2,4-DB	10.002	9.526	2966.1E6	458.6E6	733.093	742.156
14) T	DINOSEB	11.156	9.851	11510.6E6	2641.3E6	744.199	743.490
15) T	Picloram	10.981	10.832	21585.9E6	4511.5E6	727.354	701.387
16) T	DCPA	11.423	10.780	18078.6E6	4590.9E6	750.024	743.762

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060822\
Data File : PS019661.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jun 2022 02:09
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 09 03:45:03 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052422.M
Quant Title : 8080.M
QLast Update : Wed May 25 02:05:34 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

