

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060920\
 Data File : PS010743.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Jun 2020 00:14
 Operator : DD\AJ
 Sample : L2951-02
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 239988

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 10 03:20:26 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
 Quant Title : 8080.M
 QLast Update : Fri May 29 15:43:50 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.229	6.921	317.6E6	235.7E6	401.543	432.203
Target Compounds							
1) T	Dalapon	2.995	2.585	33020498	115.0E6	33.585	157.140 #
2) T	3,5-DICHL...	0.000	6.101	0	101.3E6	N.D.	123.595
3) T	4-Nitroph...	7.109	0.000	6232694	0	16.812	N.D. #
5) T	DICAMBA	7.409	7.055	7119776	11352937	2.264	5.044 #
6) T	MCPD	7.642	0.000	6298725	0	2.550	N.D. #
7) T	MCPA	7.686	7.460	62167171	50462301	17.782	18.568
8) T	DICHLORPROP	0.000	7.740	0	7976527	N.D.	13.833
11) T	2,4,5-TP ...	0.000	8.844	0	3855534	N.D.	1.284
12) T	2,4,5-T	9.428	0.000	51483679	0	12.659	N.D. #
14) T	DINOSEB	11.170	10.066	20648299	32010937	6.391	14.879 #
15) T	Picloram	10.995	0.000	47058401	0	9.218	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060920\
Data File : PS010743.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jun 2020 00:14
Operator : DD\AJ
Sample : L2951-02
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
239988

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 10 03:20:26 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
Quant Title : 8080.M
QLast Update : Fri May 29 15:43:50 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.995 min
Delta R.T.: -0.046 min
Response: 33020498
Conc: 33.58 ng/ml

Instrument :
ECD_S
ClientSampleId :
239988




#1 Dalapon

R.T.: 2.585 min
Delta R.T.: -0.038 min
Response: 115049127
Conc: 157.14 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 6.443 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.101 min
Delta R.T.: 0.030 min
Response: 101303771
Conc: 123.59 ng/ml



#3 4-Nitrophenol


R.T.: 7.109 min
Delta R.T.: 0.049 min
Response: 6232694
Conc: 16.81 ng/ml

Instrument :
ECD_S
ClientSampleId :
239988




#3 4-Nitrophenol

R.T.: 0.000 min
Exp R.T.: 6.525 min
Response: 0
Conc: N.D.




#4 2,4-DCAA

R.T.: 7.229 min
Delta R.T.: -0.002 min
Response: 317564755
Conc: 401.54 ng/ml



#4 2,4-DCAA


R.T.: 6.921 min
Delta R.T.: 0.000 min
Response: 235730575
Conc: 432.20 ng/ml



#5 DICAMBA


R.T.: 7.409 min
Delta R.T.: -0.014 min
Response: 7119776
Conc: 2.26 ng/ml

Instrument :
ECD_S
ClientSampleId :
239988




#5 DICAMBA

R.T.: 7.055 min
Delta R.T.: -0.036 min
Response: 11352937
Conc: 5.04 ng/ml




#6 MCP

R.T.: 7.642 min
Delta R.T.: 0.043 min
Response: 6298725
Conc: 2.55 ug/ml



#6 MCP


R.T.: 0.000 min
Exp R.T.: 7.196 min
Response: 0
Conc: N.D.



#7 MCPA


R.T.: 7.686 min
Delta R.T.: -0.065 min
Response: 62167171
Conc: 17.78 ug/ml

Instrument :
ECD_S
ClientSampleId :
239988




#7 MCPA

R.T.: 7.460 min
Delta R.T.: 0.046 min
Response: 50462301
Conc: 18.57 ug/ml




#8 DICHLORPROP

R.T.: 0.000 min
Exp R.T. : 8.116 min
Response: 0
Conc: N.D.



#8 DICHLORPROP


R.T.: 7.740 min
Delta R.T.: 0.022 min
Response: 7976527
Conc: 13.83 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 0.000 min
Exp R.T.: 9.208 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
239988




#11 2,4,5-TP (SILVEX)

R.T.: 8.844 min
Delta R.T.: 0.042 min
Response: 3855534
Conc: 1.28 ng/ml




#12 2,4,5-T

R.T.: 9.428 min
Delta R.T.: -0.064 min
Response: 51483679
Conc: 12.66 ng/ml



#12 2,4,5-T


R.T.: 0.000 min
Exp R.T.: 9.181 min
Response: 0
Conc: N.D.



#14 DINOSEB


R.T.: 11.170 min
Delta R.T.: -0.037 min
Response: 20648299
Conc: 6.39 ng/ml

Instrument :
ECD_S
ClientSampleId :
239988




#14 DINOSEB

R.T.: 10.066 min
Delta R.T.: 0.044 min
Response: 32010937
Conc: 14.88 ng/ml



#15 Picloram

R.T.: 10.995 min
Delta R.T.: -0.029 min
Response: 47058401
Conc: 9.22 ng/ml



#15 Picloram

R.T.: 0.000 min
Exp R.T.: 11.002 min
Response: 0
Conc: N.D.