

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060920\
 Data File : PS010734.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jun 2020 20:47
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 10 03:17:42 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
 Quant Title : 8080.M
 QLast Update : Fri May 29 15:43:50 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.229	6.921	671.6E6	447.5E6	849.253	820.483
Target Compounds							
1) T	Dalapon	3.043	2.627	685.6E6	477.8E6	697.299	652.615
2) T	3,5-DICHL...	6.442	6.070	866.5E6	626.3E6	792.615	764.138
3) T	4-Nitroph...	7.059	6.526	296.4E6	254.0E6	799.393	766.820
5) T	DICAMBA	7.421	7.090	2513.9E6	1719.8E6	799.476	764.082
6) T	MCPD	7.592	7.189	198.2E6	146.1E6	80.230	78.376
7) T	MCPA	7.742	7.405	272.7E6	208.8E6	77.993	76.814
8) T	DICHLORPROP	8.114	7.716	603.8E6	443.5E6	796.702	769.078
9) T	2,4-D	8.343	8.010	634.0E6	549.9E6	804.233	725.227
10) T	Pentachlo...	8.655	8.441	8588.5E6	5342.7E6	805.214	752.018
11) T	2,4,5-TP ...	9.204	8.801	3449.5E6	2323.6E6	806.557	773.665
12) T	2,4,5-T	9.488	9.180	3304.4E6	2337.6E6	812.496	778.970
13) T	2,4-DB	10.041	9.688	449.2E6	321.3E6	812.352	791.042
14) T	DINOSEB	11.204	10.021	2643.3E6	1705.9E6	818.069	792.888
15) T	Picloram	11.021	11.002	3811.4E6	2925.7E6	746.574	709.088
16) T	DCPA	11.469	10.949	4325.9E6	2912.2E6	807.369	785.351

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060920\
Data File : PS010734.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jun 2020 20:47
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 10 03:17:42 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
Quant Title : 8080.M
QLast Update : Fri May 29 15:43:50 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

