

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_S\Data\PS060921\  
 Data File : PS015318.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 09 Jun 2021 14:18  
 Operator : DD\AJ  
 Sample : HSTDCCC750  
 Misc :  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 ECD\_S  
 ClientSampleId :  
 HSTDCCC750

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Jun 09 14:49:18 2021  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_S\Method\PS052821.M  
 Quant Title : 8080.M  
 QLast Update : Fri May 28 14:34:22 2021  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 2 µl  
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2  
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

|                             | Compound     | RT#1   | RT#2   | Resp#1    | Resp#2   | ng/ml   | ng/ml   |
|-----------------------------|--------------|--------|--------|-----------|----------|---------|---------|
| -----                       |              |        |        |           |          |         |         |
| System Monitoring Compounds |              |        |        |           |          |         |         |
| 4) S                        | 2,4-DCAA     | 7.224  | 6.783  | 1735.6E6  | 386.7E6  | 688.635 | 721.837 |
| Target Compounds            |              |        |        |           |          |         |         |
| 1) T                        | Dalapon      | 3.038  | 2.523  | 1635.9E6  | 432.9E6  | 622.406 | 661.499 |
| 2) T                        | 3,5-DICHL... | 6.436  | 5.970  | 2173.7E6  | 542.7E6  | 630.353 | 671.752 |
| 3) T                        | 4-Nitroph... | 7.050  | 6.400  | 870.6E6   | 208.2E6  | 617.598 | 654.282 |
| 5) T                        | DICAMBA      | 7.417  | 6.949  | 6343.8E6  | 1540.1E6 | 665.548 | 695.952 |
| 6) T                        | MCPD         | 7.588  | 7.046  | 501.7E6   | 123.5E6  | 69.722  | 69.896  |
| 7) T                        | MCPA         | 7.737  | 7.259  | 684.2E6   | 180.2E6  | 68.171  | 69.338  |
| 8) T                        | DICHLORPROP  | 8.112  | 7.570  | 1521.8E6  | 373.5E6  | 660.760 | 687.515 |
| 9) T                        | 2,4-D        | 8.339  | 7.861  | 1683.2E6  | 441.6E6  | 660.568 | 685.690 |
| 10) T                       | Pentachlo... | 8.653  | 8.292  | 21491.0E6 | 5148.8E6 | 663.901 | 699.453 |
| 11) T                       | 2,4,5-TP ... | 9.203  | 8.653  | 8388.9E6  | 2051.1E6 | 675.924 | 706.798 |
| 12) T                       | 2,4,5-T      | 9.486  | 9.030  | 8206.5E6  | 2071.2E6 | 683.253 | 715.451 |
| 13) T                       | 2,4-DB       | 10.035 | 9.537  | 1790.1E6  | 271.1E6  | 702.968 | 702.899 |
| 14) T                       | DINOSEB      | 11.200 | 9.866  | 5986.8E6  | 1478.8E6 | 749.324 | 809.798 |
| 15) T                       | Picloram     | 11.014 | 10.843 | 11005.5E6 | 2887.6E6 | 664.312 | 695.311 |
| 16) T                       | DCPA         | 11.469 | 10.798 | 11782.1E6 | 3123.9E6 | 698.730 | 724.635 |
| -----                       |              |        |        |           |          |         |         |

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_S\Data\PS060921\  
Data File : PS015318.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 09 Jun 2021 14:18  
Operator : DD\AJ  
Sample : HSTDCCC750  
Misc :  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
ECD\_S  
ClientSampleId :  
HSTDCCC750

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Jun 09 14:49:18 2021  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_S\Method\PS052821.M  
Quant Title : 8080.M  
QLast Update : Fri May 28 14:34:22 2021  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 2 µl  
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2  
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

