

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060922\
 Data File : PS019691.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jun 2022 23:33
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 10 04:14:09 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052422.M
 Quant Title : 8080.M
 QLast Update : Wed May 25 02:05:34 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.197	6.782	1749.8E6	592.9E6	729.338	765.133
Target Compounds							
1) T	Dalapon	3.012	2.524	3239.0E6	684.5E6	639.008	682.194
2) T	3,5-DICHL...	6.411	5.962	2162.2E6	871.5E6	684.111	727.093
3) T	4-Nitroph...	7.030	6.400	728.7E6	302.0E6	742.545	705.577
5) T	DICAMBA	7.387	6.946	6500.4E6	2523.5E6	684.411	738.189
6) T	MCPD	7.556	7.043	516.4E6	198.7E6	68.797	73.412
7) T	MCPA	7.705	7.255	729.4E6	275.8E6	67.511	71.764
8) T	DICHLORPROP	8.078	7.566	1713.3E6	642.5E6	669.093	727.368
9) T	2,4-D	8.308	7.858	1956.1E6	770.4E6	684.079	722.313
10) T	Pentachlo...	8.614	8.282	23560.0E6	8602.1E6	681.144	737.980
11) T	2,4,5-TP ...	9.164	8.643	10925.5E6	3720.8E6	695.847	740.136
12) T	2,4,5-T	9.450	9.021	12957.1E6	3474.1E6	688.489	738.673
13) T	2,4-DB	10.001	9.526	2906.8E6	454.1E6	718.442	734.789
14) T	DINOSEB	11.155	9.850	11465.0E6	2632.2E6	741.249	740.929
15) T	Picloram	10.981	10.832	21030.5E6	4344.4E6	708.641	675.407
16) T	DCPA	11.423	10.779	17987.2E6	4559.9E6	746.233	738.739

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS060922\
Data File : PS019691.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jun 2022 23:33
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 10 04:14:09 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052422.M
Quant Title : 8080.M
QLast Update : Wed May 25 02:05:34 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

