

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061020\
 Data File : PS010760.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Jun 2020 12:20
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 10 17:51:15 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
 Quant Title : 8080.M
 QLast Update : Fri May 29 15:43:50 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.229	6.921	637.4E6	418.6E6	805.944	767.466
Target Compounds							
1) T	Dalapon	3.040	2.623	683.3E6	474.4E6	695.022	647.976
2) T	3,5-DICHL...	6.442	6.070	822.5E6	585.0E6	752.328	713.734
3) T	4-Nitroph...	7.059	6.526	279.0E6	235.2E6	752.584	710.100
5) T	DICAMBA	7.422	7.090	2394.9E6	1621.5E6	761.610	720.412
6) T	MCPD	7.592	7.189	187.6E6	136.0E6	75.934	72.974
7) T	MCPA	7.742	7.405	260.8E6	197.7E6	74.592	72.743
8) T	DICHLORPROP	8.114	7.716	574.2E6	414.8E6	757.630	719.332
9) T	2,4-D	8.343	8.010	601.5E6	515.0E6	763.078	679.196
10) T	Pentachlo...	8.655	8.441	8242.7E6	4998.8E6	772.794	703.615
11) T	2,4,5-TP ...	9.204	8.801	3302.0E6	2196.0E6	772.061	731.181
12) T	2,4,5-T	9.489	9.180	3169.2E6	2207.9E6	779.234	735.747
13) T	2,4-DB	10.041	9.687	426.8E6	303.5E6	771.882	747.240
14) T	DINOSEB	11.205	10.021	2463.2E6	1606.7E6	762.355	746.799
15) T	Picloram	11.021	11.002	3808.9E6	2889.2E6	746.092	700.250
16) T	DCPA	11.469	10.949	4089.7E6	2762.8E6	763.286	745.045

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061020\
Data File : PS010760.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jun 2020 12:20
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 10 17:51:15 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
Quant Title : 8080.M
QLast Update : Fri May 29 15:43:50 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

