

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061021\
 Data File : PS015357.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Jun 2021 23:04
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 11 04:27:08 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052821.M
 Quant Title : 8080.M
 QLast Update : Fri May 28 14:34:22 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.224	6.783	1666.1E6	376.2E6	661.063	702.099
Target Compounds							
1) T	Dalapon	3.038	2.521	1532.6E6	441.1E6	583.113	673.945m
2) T	3,5-DICHL...	6.436	5.969	2082.1E6	527.0E6	603.784	652.386
3) T	4-Nitroph...	7.050	6.399	832.1E6	204.4E6	590.242	642.264
5) T	DICAMBA	7.417	6.948	6137.5E6	1497.0E6	643.903	676.492
6) T	MCPD	7.588	7.045	481.3E6	121.5E6	66.876	68.760
7) T	MCPA	7.736	7.258	661.8E6	175.5E6	65.938	67.539
8) T	DICHLORPROP	8.111	7.569	1474.8E6	367.7E6	640.372	676.847
9) T	2,4-D	8.338	7.860	1645.3E6	433.5E6	645.711	673.117
10) T	Pentachlo...	8.652	8.290	21159.8E6	5018.8E6	653.670	681.790
11) T	2,4,5-TP ...	9.202	8.651	8255.9E6	2002.0E6	665.209	689.868
12) T	2,4,5-T	9.485	9.029	8152.7E6	2035.0E6	678.777	702.947
13) T	2,4-DB	10.035	9.535	1747.9E6	267.7E6	686.384	693.877
14) T	DINOSEB	11.199	9.864	5838.2E6	1440.7E6	730.724	788.934
15) T	Picloram	11.013	10.842	10988.0E6	2877.2E6	663.252	692.814
16) T	DCPA	11.467	10.797	11607.2E6	3065.2E6	688.362	711.008

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061021\
Data File : PS015357.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jun 2021 23:04
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 11 04:27:08 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052821.M
Quant Title : 8080.M
QLast Update : Fri May 28 14:34:22 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

