

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061022\
 Data File : PS019694.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Jun 2022 09:24
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 10 14:33:11 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052422.M
 Quant Title : 8080.M
 QLast Update : Wed May 25 02:05:34 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.198	6.781	1817.1E6	606.9E6	757.407	783.202
Target Compounds							
1) T	Dalapon	3.014	2.526	3365.7E6	689.7E6	664.013	687.290
2) T	3,5-DICHL...	6.412	5.962	2247.3E6	892.4E6	711.025	744.454
3) T	4-Nitroph...	7.030	6.400	744.1E6	305.7E6	758.228	714.300
5) T	DICAMBA	7.387	6.945	6792.1E6	2582.0E6	715.129	755.312
6) T	MCPD	7.556	7.043	543.0E6	203.4E6	72.346	75.153
7) T	MCPA	7.705	7.255	763.1E6	282.9E6	70.635	73.612
8) T	DICHLORPROP	8.078	7.565	1792.2E6	658.4E6	699.896	745.323
9) T	2,4-D	8.309	7.857	2038.2E6	786.9E6	712.781	737.735
10) T	Pentachlo...	8.614	8.282	24618.1E6	8797.7E6	711.734	754.767
11) T	2,4,5-TP ...	9.164	8.643	11067.3E6	3805.2E6	704.881	756.926
12) T	2,4,5-T	9.450	9.021	13133.0E6	3550.5E6	697.834	754.919
13) T	2,4-DB	10.002	9.525	2969.8E6	463.6E6	733.998	750.177
14) T	DINOSEB	11.155	9.850	11889.4E6	2684.8E6	768.685	755.714
15) T	Picloram	10.981	10.831	22835.8E6	4597.8E6	769.472	714.810
16) T	DCPA	11.424	10.779	18607.7E6	4660.5E6	771.975	755.038

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061022\
Data File : PS019694.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jun 2022 09:24
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 10 14:33:11 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052422.M
Quant Title : 8080.M
QLast Update : Wed May 25 02:05:34 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

