

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061121\
 Data File : PS015418.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2021 23:36
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 6/14/2021 12:27:16 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 04:34:59 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052821.M
 Quant Title : 8080.M
 QLast Update : Fri May 28 14:34:22 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.221	6.782	1859.8E6	388.8E6	737.891	725.728
Target Compounds							
1) T	Dalapon	3.036	2.519	1539.1E6	405.2E6	585.568	619.147
2) T	3,5-DICHL...	6.433	5.968	2373.2E6	543.6E6	688.210	672.928
3) T	4-Nitroph...	7.047	6.399	951.2E6	211.3E6	674.785	664.167
5) T	DICAMBA	7.414	6.947	6803.5E6	1543.8E6	713.772m	697.611
6) T	MCPD	7.584	7.045	540.0E6	125.0E6	75.036	70.728
7) T	MCPA	7.733	7.257	634.5E6	181.5E6	63.222	69.836
8) T	DICHLORPROP	8.109	7.568	1412.4E6	379.1E6	613.265	697.852
9) T	2,4-D	8.336	7.859	1562.0E6	445.6E6	613.009	691.968
10) T	Pentachlo...	8.649	8.290	20283.0E6	5175.0E6	626.584	703.010
11) T	2,4,5-TP ...	9.200	8.651	7859.4E6	2067.5E6	633.257	712.457
12) T	2,4,5-T	9.483	9.028	7652.5E6	2083.6E6	637.132	719.731
13) T	2,4-DB	10.033	9.534	1589.9E6	277.2E6	624.324	718.610
14) T	DINOSEB	11.198	9.863	5495.5E6	1476.4E6	687.837	808.490
15) T	Picloram	11.012	10.840	9774.0E6	2895.9E6	589.974	697.315
16) T	DCPA	11.466	10.796	11141.9E6	3179.6E6	660.769	737.545

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061121\
Data File : PS015418.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2021 23:36
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Ankita
6/14/2021 12:27:16 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 12 04:34:59 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052821.M
Quant Title : 8080.M
QLast Update : Fri May 28 14:34:22 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

