

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061121\
 Data File : PS015443.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jun 2021 14:55
 Operator : DD\AJ
 Sample : PB136979BS
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB136979BS

Manual Integrations
 APPROVED

Ankita
 6/14/2021 12:27:31 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 14 02:20:26 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052821.M
 Quant Title : 8080.M
 QLast Update : Fri May 28 14:34:22 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.228	6.782	1388.3E6	275.8E6	550.825	514.710
Target Compounds							
1) T	Dalapon	3.042	2.520	1427.7E6	333.8E6	543.169	510.002m
2) T	3,5-DICHL...	6.440	5.968	1980.7E6	408.8E6	574.390	506.007
3) T	4-Nitroph...	7.055	6.399	677.3E6	158.0E6	480.463	496.477
5) T	DICAMBA	7.421	6.947	5268.8E6	1097.1E6	552.765	495.748
6) T	MCPD	7.590	7.042	449.1E6	88816583	62.406	50.261
7) T	MCPA	7.739	7.254	574.0E6	141.9E6	57.190	54.604
8) T	DICHLORPROP	8.115	7.568	1365.4E6	295.1E6	592.862	543.205
9) T	2,4-D	8.343	7.858	1659.5E6	382.6E6	651.256	594.151m
10) T	Pentachlo...	8.657	8.289	19804.0E6	4052.3E6	611.785	550.502
11) T	2,4,5-TP ...	9.207	8.650	8083.6E6	1669.5E6	651.323	575.308m
12) T	2,4,5-T	9.490	9.027	7186.9E6	1568.6E6	598.366	541.822m
13) T	2,4-DB	10.044	9.534	1040.6E6	215.0E6	408.651	557.407 #
14) T	DINOSEB	11.204	9.862	5470.6E6	1072.1E6	684.722	587.100
15) T	Picloram	11.018	10.839	8128.1E6	1842.5E6	490.627	443.675
16) T	DCPA	11.472	10.795	8729.3E6	2016.9E6	517.689	467.844

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061121\
Data File : PS015443.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jun 2021 14:55
Operator : DD\AJ
Sample : PB136979BS
Misc :
ALS Vial : 44 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB136979BS

Manual Integrations
APPROVED

Ankita
6/14/2021 12:27:31 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 14 02:20:26 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052821.M
Quant Title : 8080.M
QLast Update : Fri May 28 14:34:22 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

