

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061121\
 Data File : PS015451.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jun 2021 18:08
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 6/17/2021 9:32:23 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 14 02:22:56 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052821.M
 Quant Title : 8080.M
 QLast Update : Fri May 28 14:34:22 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.228	6.781	2165.8E6	445.2E6	859.330	830.947
Target Compounds							
1) T	Dalapon	3.043	2.521	2048.3E6	447.1E6	779.300	683.215
2) T	3,5-DICHL...	6.440	5.968	2713.6E6	620.0E6	786.918	767.509
3) T	4-Nitroph...	7.055	6.398	1062.9E6	239.9E6	753.993	754.060
5) T	DICAMBA	7.422	6.946	8004.2E6	1759.3E6	839.745	795.028
6) T	MCPD	7.593	7.045	643.1E6	144.6E6	89.366	81.806
7) T	MCPA	7.741	7.258	871.1E6	208.3E6	86.794	80.170
8) T	DICHLORPROP	8.115	7.567	1930.6E6	433.0E6	838.240	797.006
9) T	2,4-D	8.343	7.857	2118.9E6	509.0E6	831.573	790.363m
10) T	Pentachlo...	8.657	8.288	27103.2E6	5746.9E6	837.273	780.707
11) T	2,4,5-TP ...	9.207	8.649	10750.4E6	2342.9E6	866.197	807.347m
12) T	2,4,5-T	9.489	9.026	10482.5E6	2329.0E6	872.748	804.507
13) T	2,4-DB	10.039	9.533	2277.9E6	316.2E6	894.494	819.669
14) T	DINOSEB	11.205	9.861	7720.8E6	1657.0E6	966.361	907.366
15) T	Picloram	11.018	10.840	13244.8E6	3121.2E6	799.479	751.575
16) T	DCPA	11.472	10.795	15477.0E6	3585.0E6	917.859	831.592

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061121\
Data File : PS015451.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jun 2021 18:08
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Ankita
6/17/2021 9:32:23 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 14 02:22:56 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052821.M
Quant Title : 8080.M
QLast Update : Fri May 28 14:34:22 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

