

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061121\
 Data File : PS015463.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jun 2021 23:22
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 6/17/2021 9:32:30 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 14 02:27:14 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052821.M
 Quant Title : 8080.M
 QLast Update : Fri May 28 14:34:22 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.228	6.781	2188.7E6	446.6E6	868.393	833.591
Target Compounds							
1) T	Dalapon	3.044	2.520	2017.0E6	443.8E6	767.385	678.157
2) T	3,5-DICHL...	6.440	5.967	2743.5E6	622.4E6	795.575	770.418
3) T	4-Nitroph...	7.054	6.398	1065.0E6	239.9E6	755.487	754.126
5) T	DICAMBA	7.421	6.946	8083.3E6	1762.6E6	848.035	796.507
6) T	MCPD	7.592	7.045	648.8E6	144.4E6	90.152	81.741
7) T	MCPA	7.742	7.258	877.1E6	208.4E6	87.394	80.174
8) T	DICHLORPROP	8.115	7.567	1944.1E6	433.6E6	844.114	798.060
9) T	2,4-D	8.343	7.857	2128.1E6	510.7E6	835.157	792.945m
10) T	Pentachlo...	8.657	8.288	27395.4E6	5773.3E6	846.299	784.295
11) T	2,4,5-TP ...	9.207	8.649	10827.1E6	2344.0E6	872.379	807.742m
12) T	2,4,5-T	9.489	9.026	10559.8E6	2332.5E6	879.186	805.682
13) T	2,4-DB	10.040	9.532	2276.0E6	315.3E6	893.772	817.314
14) T	DINOSEB	11.204	9.861	7752.3E6	1654.3E6	970.307	905.911m
15) T	Picloram	11.019	10.839	13349.5E6	3135.6E6	805.798	755.035
16) T	DCPA	11.471	10.794	15621.4E6	3587.4E6	926.419	832.155

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061121\
Data File : PS015463.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jun 2021 23:22
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Ankita
6/17/2021 9:32:30 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 14 02:27:14 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052821.M
Quant Title : 8080.M
QLast Update : Fri May 28 14:34:22 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

