

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061220\
 Data File : PS010794.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jun 2020 12:10
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 14:41:25 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
 Quant Title : 8080.M
 QLast Update : Fri May 29 15:43:50 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.229	6.920	649.7E6	421.8E6	821.546	773.402
Target Compounds							
1) T	Dalapon	3.041	2.625	710.8E6	483.1E6	722.965	659.775
2) T	3,5-DICHL...	6.442	6.069	834.4E6	590.1E6	763.213	719.956
3) T	4-Nitroph...	7.059	6.525	289.3E6	241.8E6	780.425	729.884
5) T	DICAMBA	7.422	7.090	2423.2E6	1623.8E6	770.615	721.432
6) T	MCPD	7.591	7.188	189.8E6	136.0E6	76.828	72.957
7) T	MCPA	7.742	7.404	261.8E6	196.1E6	74.894	72.153
8) T	DICHLORPROP	8.114	7.715	584.4E6	417.5E6	771.041	724.015
9) T	2,4-D	8.343	8.009	623.3E6	522.9E6	790.669	689.628
10) T	Pentachlo...	8.656	8.441	8449.1E6	4965.7E6	792.143	698.954
11) T	2,4,5-TP ...	9.204	8.801	3361.7E6	2209.5E6	786.036	735.686
12) T	2,4,5-T	9.489	9.180	3252.7E6	2238.3E6	799.788	745.866
13) T	2,4-DB	10.041	9.685	442.2E6	309.9E6	799.760	762.780
14) T	DINOSEB	11.204	10.020	2499.5E6	1626.4E6	773.585	755.945
15) T	Picloram	11.021	11.001	4163.9E6	3079.6E6	815.628	746.398
16) T	DCPA	11.468	10.949	4163.4E6	2775.0E6	777.053	748.337

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061220\
Data File : PS010794.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jun 2020 12:10
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 12 14:41:25 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS052920.M
Quant Title : 8080.M
QLast Update : Fri May 29 15:43:50 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

