

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_S\Data\PS061421\  
 Data File : PS015568.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 14 Jun 2021 20:36  
 Operator : DD\AJ  
 Sample : HSTDCCC750  
 Misc :  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 ECD\_S  
 ClientSampleId :  
 HSTDCCC750

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Jun 15 09:07:01 2021  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_S\Method\PS061421.M  
 Quant Title : 8080.M  
 QLast Update : Tue Jun 15 08:41:47 2021  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 2 µl  
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2  
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

|                             | Compound     | RT#1   | RT#2   | Resp#1    | Resp#2   | ng/ml   | ng/ml   |
|-----------------------------|--------------|--------|--------|-----------|----------|---------|---------|
| -----                       |              |        |        |           |          |         |         |
| System Monitoring Compounds |              |        |        |           |          |         |         |
| 4) S                        | 2,4-DCAA     | 7.229  | 6.780  | 2554.8E6  | 407.9E6  | 918.619 | 757.334 |
| Target Compounds            |              |        |        |           |          |         |         |
| 1) T                        | Dalapon      | 3.048  | 2.524  | 2661.4E6  | 444.1E6  | 928.060 | 721.182 |
| 2) T                        | 3,5-DICHL... | 6.441  | 5.967  | 3231.0E6  | 574.8E6  | 847.487 | 708.895 |
| 3) T                        | 4-Nitroph... | 7.056  | 6.398  | 1288.4E6  | 220.3E6  | 851.866 | 696.416 |
| 5) T                        | DICAMBA      | 7.422  | 6.945  | 9417.3E6  | 1620.0E6 | 879.092 | 724.640 |
| 6) T                        | MCPD         | 7.593  | 7.043  | 757.1E6   | 130.8E6  | 93.856  | 73.744  |
| 7) T                        | MCPA         | 7.741  | 7.256  | 1007.2E6  | 190.5E6  | 90.012  | 71.564  |
| 8) T                        | DICHLORPROP  | 8.116  | 7.566  | 2285.5E6  | 394.1E6  | 876.463 | 715.233 |
| 9) T                        | 2,4-D        | 8.344  | 7.856  | 2513.3E6  | 465.2E6  | 866.733 | 713.443 |
| 10) T                       | Pentachlo... | 8.657  | 8.287  | 31922.3E6 | 5437.3E6 | 872.091 | 727.606 |
| 11) T                       | 2,4,5-TP ... | 9.207  | 8.648  | 12568.0E6 | 2164.3E6 | 868.678 | 731.889 |
| 12) T                       | 2,4,5-T      | 9.490  | 9.025  | 12217.0E6 | 2153.5E6 | 861.943 | 731.023 |
| 13) T                       | 2,4-DB       | 10.040 | 9.531  | 2647.5E6  | 284.2E6  | 838.078 | 717.940 |
| 14) T                       | DINOSEB      | 11.205 | 9.860  | 8762.1E6  | 1475.9E6 | 935.516 | 820.843 |
| 15) T                       | Picloram     | 11.019 | 10.837 | 16214.3E6 | 2999.1E6 | 844.085 | 723.633 |
| 16) T                       | DCPA         | 11.472 | 10.792 | 17931.2E6 | 3300.5E6 | 861.592 | 732.665 |
| -----                       |              |        |        |           |          |         |         |

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_S\Data\PS061421\  
Data File : PS015568.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 14 Jun 2021 20:36  
Operator : DD\AJ  
Sample : HSTDCCC750  
Misc :  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
ECD\_S  
ClientSampleId :  
HSTDCCC750

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Jun 15 09:07:01 2021  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_S\Method\PS061421.M  
Quant Title : 8080.M  
QLast Update : Tue Jun 15 08:41:47 2021  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 2 µl  
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2  
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

