

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061421\
 Data File : PS015592.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Jun 2021 06:16
 Operator : DD\AJ
 Sample : HSTDCCC750 (Sig #1); PSTDCCC050 (Sig #2)
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 15 09:13:24 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061421.M
 Quant Title : 8080.M
 QLast Update : Tue Jun 15 08:41:47 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.231	6.780	2309.9E6	408.6E6	830.572	758.717
Target Compounds							
1) T	Dalapon	3.046	2.523	2682.2E6	453.7E6	935.324	736.865
2) T	3,5-DICHL...	6.441	5.967	3115.9E6	576.7E6	817.314	711.284
3) T	4-Nitroph...	7.057	6.398	1167.1E6	222.7E6	771.687	703.804
5) T	DICAMBA	7.423	6.945	8447.1E6	1623.5E6	788.522	726.186
6) T	MCPD	7.594	7.043	654.8E6	131.4E6	81.183	74.098
7) T	MCPA	7.743	7.256	878.4E6	190.4E6	78.502	71.510
8) T	DICHLORPROP	8.118	7.566	1965.4E6	395.5E6	753.702	717.891
9) T	2,4-D	8.345	7.856	2151.8E6	467.1E6	742.083	716.441
10) T	Pentachlo...	8.659	8.287	27575.7E6	5428.4E6	753.346	726.412
11) T	2,4,5-TP ...	9.208	8.648	10583.9E6	2164.6E6	731.541	731.998
12) T	2,4,5-T	9.491	9.025	10233.9E6	2172.6E6	722.030	737.515
13) T	2,4-DB	10.042	9.531	2108.1E6	284.3E6	667.309	718.365
14) T	DINOSEB	11.206	9.859	7334.7E6	1538.6E6	783.113	855.718
15) T	Picloram	11.020	10.836	13497.6E6	3051.4E6	702.659	736.240
16) T	DCPA	11.473	10.792	14856.6E6	3301.1E6	713.861	732.791

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061421\
Data File : PS015592.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Jun 2021 06:16
Operator : DD\AJ
Sample : HSTDCCC750 (Sig #1); PSTDCCC050 (Sig #2)
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 15 09:13:24 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061421.M
Quant Title : 8080.M
QLast Update : Tue Jun 15 08:41:47 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

