

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061521\
 Data File : PS015602.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Jun 2021 16:44
 Operator : DD\AJ
 Sample : HSTDICC750
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 15 18:11:34 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061521.M
 Quant Title : 8080.M
 QLast Update : Tue Jun 15 17:55:22 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.193	6.770	1287.7E6	388.1E6	750.000	750.000
Target Compounds							
1) T	Dalapon	3.021	2.522	1433.3E6	457.7E6	682.500	682.500
2) T	3,5-DICHL...	6.410	5.960	1622.4E6	544.4E6	697.500	697.500
3) T	4-Nitroph...	7.020	6.389	642.3E6	214.9E6	682.500	682.500
5) T	DICAMBA	7.385	6.934	4748.8E6	1521.7E6	705.000	705.000
6) T	MCPD	7.556	7.032	400.5E6	121.3E6	70.500	70.500
7) T	MCPA	7.705	7.244	564.2E6	181.4E6	69.750	69.750
8) T	DICHLORPROP	8.078	7.554	1153.1E6	374.2E6	705.000	705.000
9) T	2,4-D	8.306	7.845	1286.2E6	450.2E6	705.000	705.000
10) T	Pentachlo...	8.617	8.275	16024.0E6	5164.9E6	712.500	712.500
11) T	2,4,5-TP ...	9.169	8.636	6504.2E6	2041.7E6	712.500	712.500
12) T	2,4,5-T	9.451	9.013	6582.9E6	2092.2E6	712.500	712.500
13) T	2,4-DB	10.001	9.519	1535.8E6	275.4E6	712.500	712.500
14) T	DINOSEB	11.163	9.847	4787.9E6	1481.6E6	705.000	705.000
15) T	Picloram	10.978	10.824	10391.2E6	3180.2E6	712.500	712.500
16) T	DCPA	11.430	10.779	9941.5E6	3131.6E6	720.000	720.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061521\
Data File : PS015602.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Jun 2021 16:44
Operator : DD\AJ
Sample : HSTDICC750
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 15 18:11:34 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061521.M
Quant Title : 8080.M
QLast Update : Tue Jun 15 17:55:22 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

