

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061521\
 Data File : PS015692.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Jun 2021 06:09
 Operator : DD\AJ
 Sample : M2613-14MSD (Sig #1); M2613-14MS (Sig #2)
 Misc :
 ALS Vial : 59 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 18-B12(86-90)MSD

Manual Integrations
 APPROVED

Ankita
 6/17/2021 9:02:47 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 17 06:52:38 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061521.M
 Quant Title : 8080.M
 QLast Update : Tue Jun 15 18:15:49 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.191	6.769	755.7E6	236.6E6	454.846	469.585m
Target Compounds							
1) T	Dalapon	3.017	2.521	898.8E6	294.0E6	445.313	452.728m
2) T	3,5-DICHL...	6.407	5.959	1230.5E6	377.0E6	548.223	499.844
5) T	DICAMBA	7.383	6.933	3231.8E6	1029.5E6	502.689	496.165
6) T	MCP	7.551	7.029	286.6E6	87035579	52.766	52.218
7) T	MCPA	7.700	7.240	371.6E6	129.4E6	47.467	50.425m
8) T	DICHLORPROP	8.076	7.554	847.0E6	321.8E6	534.958	621.624
9) T	2,4-D	8.303	7.843	1034.0E6	356.7E6	590.006	575.082m
10) T	Pentachlo...	8.614	8.273	10787.3E6	3516.9E6	499.797	500.727
11) T	2,4,5-TP ...	9.165	8.633	4749.4E6	1995.7E6	543.339	722.874m#
12) T	2,4,5-T	9.448	9.011	4408.8E6	1545.9E6	499.390	548.100
13) T	2,4-DB	10.000	9.517	531.3E6	198.4E6	259.748m	531.279 #
14) T	DINOSEB	11.159	9.845	2519.8E6	870.7E6	389.068	432.293
15) T	Picloram	10.974	10.822	5096.3E6	1877.8E6	372.783	446.506
16) T	DCPA	11.426	10.776	4707.5E6	1897.4E6	355.916	451.302 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061521\
Data File : PS015692.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Jun 2021 06:09
Operator : DD\AJ
Sample : M2613-14MSD (Sig #1); M2613-14MS (Sig #2)
Misc :
ALS Vial : 59 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
18-B12(86-90)MSD

Manual Integrations
APPROVED

Ankita
6/17/2021 9:02:47 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 17 06:52:38 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061521.M
Quant Title : 8080.M
QLast Update : Tue Jun 15 18:15:49 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

