

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061625\
 Data File : PS030683.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Jun 2025 10:44
 Operator : AR\AJ
 Sample : HSTDICC750
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC750

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 06/17/2025
 Supervised By :mohammad ahmed 06/18/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 16 11:24:44 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061625.M
 Quant Title : 8080.M
 QLast Update : Mon Jun 16 11:23:46 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.350	7.768	2613.6E6	754.7E6	737.966m	750.000
Target Compounds							
1) T	Dalapon	2.709	2.709	4039.6E6	1799.3E6	682.500	682.500
2) T	3,5-DICHL...	6.509	6.713	3548.4E6	1052.9E6	697.500	697.500
3) T	4-Nitroph...	7.151	7.300	1040.1E6	1102.6E6	682.500	682.500
5) T	DICAMBA	7.542	7.971	10471.5E6	4437.7E6	705.000	705.000
6) T	MCPD	7.724	8.070	662.7E6	153.4E6	70.500	70.500
7) T	MCPA	7.874	8.319	799.8E6	219.4E6	69.750	69.750
8) T	DICHLORPROP	8.256	8.690	2360.0E6	1010.1E6	705.000	705.000
9) T	2,4-D	8.488	9.026	2235.4E6	1103.9E6	705.000	705.000
10) T	Pentachlo...	8.796	9.551	36925.2E6	26257.6E6	712.500	712.500
11) T	2,4,5-TP ...	9.377	9.932	13586.7E6	9730.6E6	712.500	712.500
12) T	2,4,5-T	9.671	10.358	11366.1E6	9338.5E6	712.500	712.500
13) T	2,4-DB	10.250	10.925	1661.3E6	793.3E6	712.500	712.500
14) T	DINOSEB	11.470	11.311	9488.0E6	7329.0E6	705.000	705.000
15) T	Picloram	11.278	12.419	11367.4E6	16595.5E6	712.500	712.500
16) T	DCPA	11.767	12.355	17851.5E6	14761.9E6	720.000	720.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061625\
Data File : PS030683.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Jun 2025 10:44
Operator : AR\AJ
Sample : HSTDICC750
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 06/17/2025
Supervised By : mohammad ahmed 06/18/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 16 11:24:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061625.M
Quant Title : 8080.M
QLast Update : Mon Jun 16 11:23:46 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

