

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061625\
 Data File : PS030686.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Jun 2025 11:57
 Operator : AR\AJ
 Sample : HSTDICV750
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 ICVPS061625

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 06/17/2025
 Supervised By :mohammad ahmed 06/18/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 16 12:24:52 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061625.M
 Quant Title : 8080.M
 QLast Update : Mon Jun 16 11:52:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.350	7.767	2558.2E6	759.8E6	680.238m	703.715
Target Compounds							
1) T	Dalapon	2.708	2.709	4014.4E6	1804.1E6	641.812	641.087
2) T	3,5-DICHL...	6.510	6.713	3528.6E6	1059.7E6	648.982	655.055
3) T	4-Nitroph...	7.150	7.299	1033.9E6	1107.1E6	634.193	649.407
5) T	DICAMBA	7.542	7.970	10445.4E6	4464.3E6	667.624	673.299
6) T	MCP	7.724	8.070	663.5E6	154.7E6	69.764	70.267
7) T	MCPA	7.874	8.319	800.3E6	218.4E6	67.159	68.132
8) T	DICHLORPROP	8.256	8.690	2361.7E6	1016.2E6	642.605	663.162
9) T	2,4-D	8.489	9.025	2237.6E6	1120.9E6	658.055	668.594
10) T	Pentachlo...	8.797	9.551	37024.8E6	26410.4E6	716.281	696.183
11) T	2,4,5-TP ...	9.378	9.931	13677.8E6	9797.4E6	683.657	682.524
12) T	2,4,5-T	9.671	10.357	11424.9E6	9399.0E6	687.087	681.655
13) T	2,4-DB	10.251	10.925	1676.4E6	798.8E6	678.397	673.208
14) T	DINOSEB	11.470	11.310	9565.7E6	7384.9E6	675.475	673.969
15) T	Picloram	11.279	12.419	11520.5E6	16837.8E6	690.007	702.281
16) T	DCPA	11.766	12.355	17672.3E6	14867.2E6	680.354	692.283

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS061625\
Data File : PS030686.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Jun 2025 11:57
Operator : AR\AJ
Sample : HSTDICV750
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
ICVPS061625

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 06/17/2025
Supervised By : mohammad ahmed 06/18/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 16 12:24:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061625.M
Quant Title : 8080.M
QLast Update : Mon Jun 16 11:52:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

