

Method Path : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_S\Method\
 Method File : PS061718.M
 Title : 8080.M
 Last Update : Mon Jun 18 00:27:44 2018
 Response Via : Initial Calibration

Calibration Files

750 =PS000404.D 1000 =PS000405.D 1500 =PS000406.D
 500 =PS000403.D 200 =PS000402.D

	Compound	750	1000	1500	500	200	Avg	%RSD
1) T	Dalapon	5.241	4.976	4.833	5.167	5.356	5.115 E5	4.10
2) T	3,5-DICHLOROBENZOI...	5.459	5.187	4.957	5.496	5.885	5.397 E5	6.48
3) T	4-Nitrophenol	2.354	2.224	2.126	2.402	2.648	2.351 E5	8.44
4) S	2,4-DCAA	3.771	3.588	3.434	3.839	4.201	3.767 E5	7.71
5) T	DICAMBA	1.450	1.383	1.322	1.425	1.487	1.413 E6	4.50
6) T	MCPP	1.166	1.143	1.170	1.133	1.050	1.132 E6	4.30
7) T	MCPA	1.773	1.727	1.694	1.755	1.839	1.758 E6	3.10
8) T	DICHLORPROP	3.761	3.602	3.469	3.825	4.150	3.762 E5	6.85
9) T	2,4-D	4.959	4.738	4.297	4.865	5.056	4.783 E5	6.19
10) T	Pentachlorophenol	5.332	4.953	4.509	5.555	5.900	5.250 E6	10.26
11) T	2,4,5-TP (SILVEX)	2.212	2.097	1.968	2.244	2.333	2.171 E6	6.50
12) T	2,4,5-T	2.201	2.092	1.980	2.215	2.278	2.153 E6	5.48
13) T	2,4-DB	2.979	2.878	2.804	2.970	3.090	2.944 E5	3.68
14) T	DINOSEB	1.509	1.434	1.369	1.519	1.614	1.489 E6	6.20
15) T	Picloram	3.015	2.881	2.751	2.972	2.984	2.920 E6	3.67
16) T	DCPA	2.602	2.445	2.263	2.676	2.826	2.562 E6	8.45

Signal #2 Calibration Files

750 =PS000404.D 1000 =PS000405.D 1500 =PS000406.D
 500 =PS000403.D 200 =PS000402.D

	Compound	750	1000	1500	500	200	Avg	%RSD
1) T	Dalapon	4.403	4.198	4.270	4.472	4.453	4.359 E5	2.74
2) T	3,5-DICHLOROBENZOI...	4.639	4.451	4.292	4.663	4.956	4.600 E5	5.42
3) T	4-Nitrophenol	1.960	1.873	1.804	1.984	2.177	1.960 E5	7.18
4) S	2,4-DCAA	3.450	3.271	3.174	3.441	3.680	3.403 E5	5.70
5) T	DICAMBA	1.270	1.231	1.190	1.253	1.251	1.239 E6	2.48
6) T	MCPP	9.859	9.720	9.610	9.643	9.142	9.595 E5	2.82
7) T	MCPA	1.536	1.458	1.438	0.992	1.500	1.385 E6	16.10
8) T	DICHLORPROP	3.233	3.111	3.034	3.256	3.464	3.220 E5	5.09
9) T	2,4-D	4.412	4.076	3.924	4.348	4.490	4.250 E5	5.65
10) T	Pentachlorophenol	4.841	4.537	4.191	4.961	5.123	4.731 E6	7.82
11) T	2,4,5-TP (SILVEX)	1.994	1.906	1.817	1.993	2.020	1.946 E6	4.33
12) T	2,4,5-T	1.933	1.526	1.460	1.920	1.616	1.691 E6	13.13
13) T	2,4-DB	2.535	2.430	2.396	2.478	2.539	2.475 E5	2.55
14) T	DINOSEB	1.330	1.274	1.232	1.318	1.369	1.305 E6	4.06
15) T	Picloram	2.607	2.519	2.433	2.522	2.391	2.494 E6	3.38
16) T	DCPA	2.359	2.231	2.098	2.389	2.453	2.306 E6	6.13

(#) = Out of Range